

TECHNICAL CHARACTERISTICS

| | | |
|--|--------------------------------------|------------------|
| TYPOLGY (MDD 93/42/EEC) | Class IIa Medical Decice | |
| MODEL | CLINIC PLUS SUCTION ASPIRATOR | |
| UNI EN ISO 10079-1 | HIGH VACUUM / HIGH FLOW | |
| POWER FEEDING | 230V~ / 50Hz | |
| POWER CONSUMPTION | 230 VA | |
| FUSE | F 1 x 4A L 250V | |
| MAXIMUM SUCTION PRESSURE (without jar) | -90kPa / -0.90 Bar / -675mmHg | |
| MAXIMUM SUCTION FLOW (without jar) | 60 l/min | |
| WEIGHT | 13 Kg | |
| SIZE | 460 x 850 (h) x 420 mm | |
| DUTY CYCLE | Non – Stop Operated | |
| SICILICONE TUBE SIZE | Ø 8 x 14 mm | |
| ACCURANCY OF VACUUM INDICATOR | ± 5% | |
| WORKING CONDITION | Room temperature: | 5 ÷ 35°C |
| | Room humidity percentage: | 30 ÷ 75% RH |
| | Atmospheric pressure: | 800 ÷ 1060 hPa |
| | Altitude: | 0 ÷ 2000m s.l.m. |
| CONSERVATION CONDITION AND TRASPORT | Room temperature: | -40÷ 70°C |
| | Room humidity percentage: | 10 ÷ 100% RH |
| | Atmospheric pressure: | 500 ÷ 1060 hPa |

CLEANING THE MAIN UNIT

To clean the device external parts always use a cotton cloth dampened with detergent. Don't use abrasive or solvent detergents.

ATTENTION: During cleaning make sure that liquids do not come into contact with the membrane keyboard (only in versions with footswitch and flux deviator) and adjacent areas as this would damage the component, with possible infiltration of the liquid inside the device.

The symbol



positioned in the casing near the membrane keyboard requires the reading of the user instructions before each use.



PARTICULAR CARE SHOULD BE TAKEN TO ENSURE THAT THE INTERNAL PARTS OF THE EQUIPMENT DO NOT GET IN TOUCH WITH LIQUIDS. NEVER USE LIQUIDS (e.g. DETERGENTS AND/OR SANITISING SUBSTANCES) TO CLEAN THE MAIN UNIT (ESPECIALLY NEAR THE MEMBRANE KEYBOARD) AS THEY MAY PENETRATE INSIDE THE DEVICE

During all clearing operations use protection gloves and apron (if need be, also wear a face mask and glasses) to avoid getting in contact with contaminating substances (after each utilization cycle of the machine).