



PEEL OPEN

# Cam Hospital®

### AVVERTENZE

Il prodotto è monouso, non riutilizzare. Il prodotto è di utilizzo di personale specializzato. Controllare l'integrità della confezione prima dell'uso. Controllare l'integrità dei componenti. Non usare se il prodotto appare danneggiato. Sterilità garantita per confezione integra. Dopo l'uso non disperdere nell'ambiente. Usare entro la data di scadenza.

### DESTINAZIONE D'USO

Per procedure ambulatoriali / chirurgiche

### MODALITA' D'USO

Aprire la confezione verificandone l'integrità ed estrarre il contenuto in maniera sterile. Poggiare il contenuto sul tavolo servitore o madre e aprire sterilmente il contenuto in maniera asettica, secondo tecniche procedurali come protocollo clinico in uso.

### RESPONSABILITA'

Cam Hospital declina ogni responsabilità per danni a persone e/o cose derivanti dall'uso scorretto e/o improprio del dispositivo o a seguito di utilizzo da parte di personale non adeguatamente addestrato.

### WARNINGS

The product is disposable and should not be reused. The product is to be used by qualified personnel. Package integrity should be checked before use. Check for the integrity of components. Do not use if the product looks damaged. Sterility is guaranteed provided that the package is intact. Do not discard into the environment after use. Use before expiry date.

### INTENDED USE

Outpatient/surgical procedures.

### INSTRUCTIONS FOR USE

Open the package, checking for its integrity, and take out its contents in a sterile manner. Place the contents on the instrument table or stand and open the contents in a sterile and aseptic manner, in compliance with the procedural techniques envisaged by the clinical protocol in use.

### LIABILITY

Cam Hospital disclaims any liability for damages to people and/or objects arising from the incorrect and/or improper use of the device or as a result of use by staff without proper training.



Cam Hospital S.r.l. - Italy



Cam Hospital S.r.l. - 87027 Paola (CS) - Tel. +39 0982 583609 - Fax +39 06 91280436 - www.camhospital.com

Cam Hospital S.r.l. - 87027 Paola (CS) - Tel. +39 0982 583609 - Fax +39 06 91280436 - www.camhospital.com



LOTTO V2111140  
SCAD. 2024/06  
Rev. 01 del 10/12/2020

26931  
000091



### KIT SUTURA

Suture Kit

Codice / Code ME-PI-103

#### Composizione / Composition:

- 01 TELO CAMPO / Field drape
- 01 TELO FENESTRATO / Fenestrated drape
- 01 PINZA CHIRURGICA / Surgical forceps
- 01 PINZA ANATOMICA / Anatomic forceps
- 01 PINZA PORT' AGO / Needle holder clamp
- 01 PINZA AD ANELLI / Sponge holder forceps
- 01 PINZA EMOSTATICA / Hemostatic forceps
- 01 FORBICE RETTA / Straight scissors
- 15 COMPR. DI GARZA cm 7,5x7,5x8 / Gauze swabs
- 05 TAMPONE DI GARZA ø mm 35 / Gauze ball
- 02 GUANTI MONOUSO M / Disposable gloves
- 01 SACCHETTO RIFIUTI G / Waste bag

DISPOSABLE - STERILITY GUARANTEED FOR THE COMPLETE PACKAGING