



# GIMA

PROFESSIONAL MEDICAL PRODUCTS

---

## **GB - USER MANUAL FOR KIT INSTRUMENTS**

These instruments are intended to be used by medical practitioners who are specially trained on how to use and care of them. The incorrect use, poor or inappropriate maintenance can rapidly lead to deterioration of the instruments. The first time, and after every use it is recommended to clean, dry and sterilise the instruments. The instruments must always be cleaned prior to sterilisation. For automated cleaning use only washing equipment with approved and certified detergents. For manual cleaning use approved and certified detergents, brush and running water. Always follow instructions on how to use the detergent; clean the instruments both when open and closed; rinse for 3 minutes and check that water also enters and exits the blind holes several times. Use complete-










ly demineralised water in the final rinse phase. Instruments that are not dried could suffer damage by corrosion. Always dry the instruments. After cleaning, and before sterilisation it is recommended that you treat the instruments with physiologically safe oil, especially the tips, connectors, terminals and all moving parts. Also make sure that the product does not come into contact with acids or other aggressive disinfectants that could corrode it. The recommended method of sterilisation is that of autoclave steam, the temperature of the cycle must not exceed 135°C (275°F) for a maximum of 15 minutes, to avoid damage to the product. The process of steam sterilisation must take place in accordance with EN ISO 17664. In the context of validating the sterilisation process, check the suitability of the specific measures for drying. The humidity in the container can cause the instruments to rust. Often bad, and insufficient drying, is due to the incorrect positioning of the load and the use of unsuitable types of cloths for drying. For hot air sterilisation, it is recommended to have a temperature ranging between 180° and 200°C. The instruments can also be disinfected in washer disinfectors up to a temperature of 121°C. There is no advice about maximum number of sterilisation cycles, this depends largely on the state of the product. Instruments that show signs of corrosion must be discarded immediately. Always perform a visual inspection for damage or signs of wear: sharp edges must be free of dents and with continuous edges; there should be no distortion of instruments with long parts; the instruments that are part of a larger assembly, must be checked together with other assembly components; always check the rotating movement of hinges, which must not have excessive play; always check that the locking systems are working.

#### **GIMA WARRANTY TERMS**

The Gima 12-month standard B2B warranty applies.

#### **WARNINGS**

**All serious accidents concerning the medical device supplied by us must be reported to the manufacturer and competent authority of the member state where your registered office is located.**

	<p><b>IT</b> Fabbricante <b>GB</b> Manufacturer <b>FR</b> Fabricant  <b>ES</b> Fabricante <b>PT</b> Fabricante <b>DE</b> Hersteller  <b>GR</b> Παραγωγός</p> <p style="text-align: right;"><b>SA</b> الشركة المصنعة</p>
	<p><b>IT</b> Data di fabbricazione <b>GB</b> Date of manufacture <b>FR</b> Date de fabrication  <b>ES</b> Fecha de fabricación <b>PT</b> Data de fabrico <b>DE</b> Herstellungsdatum  <b>GR</b> Ημερομηνία παραγωγής</p> <p style="text-align: right;"><b>SA</b> تاريخ التصنيع</p>
	<p><b>IT</b> Non sterile <b>GB</b> Non-sterile <b>FR</b> Pas stérile  <b>ES</b> No estéril <b>PT</b> Não estéril <b>DE</b> Nicht steril  <b>GR</b> Όχι αποστειρωμένο</p> <p style="text-align: right;"><b>SA</b> ليس معقم</p>
	<p><b>IT</b> <b>Attenzione:</b> Leggere e seguire attentamente le istruzioni (avvertenze) per l'uso  <b>GB</b> <b>Caution:</b> read instructions (warnings) carefully  <b>FR</b> <b>Attention:</b> lisez attentivement les instructions (avertissements)  <b>ES</b> <b>Precaución:</b> lea las instrucciones (advertencias) cuidadosamente  <b>PT</b> <b>Cuidado:</b> leia as instruções (avisos) cuidadosamente  <b>DE</b> <b>Achtung:</b> Anweisungen (Warnungen) sorgfältig lesen  <b>GR</b> <b>Προσοχή:</b> διαβάστε προσεκτικά τις οδηγίες (ενστάσεις)</p> <p style="text-align: right;"><b>SA</b> الحذر: قراءة التعليمات (التحذيرات) بعناية</p>
	<p><b>IT</b> Leggere le istruzioni per l'uso  <b>GB</b> Consult instructions for use  <b>FR</b> Consulter les instructions d'utilisation  <b>ES</b> Consultar las instrucciones de uso  <b>PT</b> Consulte as instruções de uso  <b>DE</b> Gebrauchsanweisung beachten  <b>GR</b> Διαβάστε προσεκτικά τις οδηγίες χρήσης</p> <p style="text-align: right;"><b>SA</b> اقرأ بدقة وحرص تعليمات الاستخدام</p>
	<p><b>IT</b> Conservare al riparo dalla luce solare  <b>GB</b> Keep away from sunlight  <b>FR</b> À conserver à l'abri de la lumière du soleil  <b>ES</b> Conservar al amparo de la luz solar  <b>PT</b> Guardar ao abrigo da luz solar  <b>DE</b> Vor Sonneneinstrahlung geschützt lagern  <b>GR</b> Κρατήστε το μακριά από ηλιακή ακτινοβολία</p> <p style="text-align: right;"><b>SA</b> يحفظ بعيدا عن أشعة الشمس</p>
	<p><b>IT</b> Conservare in luogo fresco ed asciutto  <b>GB</b> Keep in a cool, dry place  <b>FR</b> À conserver dans un endroit frais et sec  <b>ES</b> Conservar en un lugar fresco y seco  <b>PT</b> Armazenar em local fresco e seco  <b>DE</b> An einem kühlen und trockenen Ort lagern  <b>GR</b> Διατηρείται σε δροσερό και στεγνό περιβάλλον</p> <p style="text-align: right;"><b>SA</b> يحفظ في مكان بارد وجاف</p>
	<p><b>IT</b> Codice prodotto <b>GB</b> Product code <b>FR</b> Code produit  <b>ES</b> Código producto <b>PT</b> Código produto <b>DE</b> Erzeugniscode  <b>GR</b> Κωδικός προϊόντος</p> <p style="text-align: right;"><b>SA</b> كود المنتج</p>
	<p><b>IT</b> Numero di lotto <b>GB</b> Lot number <b>FR</b> Numéro de lot  <b>ES</b> Número de lote <b>PT</b> Número de lote <b>DE</b> Chargennummer  <b>GR</b> Αριθμός παρτίδας</p> <p style="text-align: right;"><b>SA</b> رقم الدفعة</p>

<b>MD</b>	<p><b>IT</b> Dispositivo medico <b>GB</b> Medical Device <b>FR</b> Dispositif médical  <b>ES</b> Producto sanitario <b>PT</b> Dispositivo médico <b>DE</b> Medizinprodukt  <b>GR</b> Ιατροτεχνολογικό προϊόν</p> <p style="text-align: right;"><b>SA</b> جهاز طبي</p>
<p><b>CE</b></p>	<p><b>IT</b> - Gli strumenti non chirurgici riutilizzabili sono da intendersi conformi al Regolamento (UE) 2017/745.  - Gli strumenti chirurgici riutilizzabili sono da intendersi conformi alla Direttiva 93/42/CEE; in accordo alla MDCG 2020-2, saranno resi conformi al Regolamento (UE) 2017/745, per cambio classe, entro maggio 2024.  <b>GB</b> - Reusable non-surgical instruments are to be considered compliant with Regulation (EU) 2017/745.  - Reusable surgical instruments are to be considered compliant with Directive 93/42/EEC; in accordance with MDCG 2020-2, they shall be made compliant with Regulation (EU) 2017/745, for change of class, by May 2024.  <b>FR</b> - Les instruments non chirurgicaux réutilisables sont destinés à être conformes au règlement (UE) 2017/745.  - Les instruments chirurgicaux réutilisables sont destinés à être conformes à la directive 93/42/CEE ; conformément au MDCG 2020-2, ils seront mis en conformité avec le règlement (UE) 2017/745, par changement de classe, d'ici mai 2024.  <b>ES</b> - Los instrumentos no quirúrgicos reutilizables deben ser conformes al Reglamento (UE) 2017/745.  - Los instrumentos quirúrgicos reutilizables deben ser conformes a la Directiva 93/42/CEE; en conformidad con la MDCG 2020-2, se adecuarán al Reglamento (UE) 2017/745 para el cambio de clase antes del mes de mayo de 2024.  <b>PT</b> - Os instrumentos não cirúrgicos reutilizáveis devem ser entendidos em conformidade com o Regulamento (UE) 2017/745.  - Os instrumentos cirúrgicos reutilizáveis devem ser entendidos em conformidade com a Diretiva 93/42/CEE; de acordo com a MDCG 2020-2, deverão estar em conformidade com o Regulamento (UE) 2017/745 devido a mudança de classe até maio de 2024.  <b>DE</b> - Die wiederverwendbaren, nicht chirurgischen Instrumente erfüllen die Anforderungen der Verordnung (EU) 2017/745.  - Die wiederverwendbaren, chirurgischen Instrumente erfüllen die Anforderungen der Richtlinie 93/42/EWG; sie werden bis Mai 2024 in Übereinstimmung mit der MDCG 2020-2 konform mit der Verordnung (EU) 2017/745 für den Wechsel der Klasse gestaltet.  <b>GR</b> - Τα επαναχρησιμοποιούμενα μη χειρουργικά εργαλεία συμμορφώνονται με τον Κανονισμό (ΕΕ) 2017/745.  - Τα επαναχρησιμοποιούμενα χειρουργικά εργαλεία συμμορφώνονται με την Οδηγία 93/42/ΕΟΚ. Σύμφωνα με το MDCG 2020-2, θα έχουν συμμορφωθεί με τον Κανονισμό (ΕΕ) 2017/745, λόγω αλλαγής κατηγορίας, μέχρι το Μάιο του 2024.</p> <p><b>SA</b> - تعتبر المعدات الغير جراحية التي يمكن إعادة استخدامها متوافقة مع توجيه لائحة الاتحاد الأوروبي 2017/745 (EU).</p> <p>تعتبر المعدات الجراحية التي يمكن إعادة استخدامها متوافقة مع توجيه لائحة الاتحاد الأوروبي CEE/42/93؛ وفقاً لمجموعة تنسيق الأجهزة الطبية MDCG 2020-2، سيتم جعلها متوافقة مع توجيه لائحة الاتحاد الأوروبي 2017/745 (EU) بموجب تغيير الفئة وفي غضون مايو 2024.</p>

**REF 26966**

Assemblato da / Assembled by / Assemblé par  
Montado por / Montado por / Zusammengesetzt von  
Συναρμολογήθηκε από / تم تجميعها من:  
**Gima S.p.A.**  
Via Marconi, 1 - 20060 Gessate (MI) Italy  
gima@gimaitaly.com - export@gimaitaly.com  
**www.gimaitaly.com**  
Made in Pakistan

**MD**

