



GIMA

PROFESSIONAL MEDICAL PRODUCTS

STAMPELLE - CRUTCHES - BÉQUILLES - KRÜCKEN MULETAS
- MULETAS - ΔΕΚΑΝΙΚΙΑ -

العكازات



All serious accidents concerning the medical device supplied by us must be reported to the manufacturer and competent authority of the member state where your registered office is located.

WARNINGS

The crutches must be used properly, holding them by the handgrips and only to support the body (weight capacity 100 Kg). Use under a doctor's direction. The crutches must be adjusted correctly according to how they need to be used and to the body of the patient following the doctor's instructions. Before use verify that the crutches are not damaged, that they are adjusted correctly and that the anti-skid rubber tips are not worn. If worn replace immediately. Be very careful when the crutches are used on wet or inclined surfaces.

ADJUSTMENT

Forearm crutches Cod. 27793

The height can be adjusted between 96 and 119 cm by pressing the pin and making it click into the right slot. The clamping screw avoids play between the two tubes and must be tightened only when the pin is inserted.

Standard crutches Cod. 27798

The height can be adjusted between 112 and 132 cm by pressing the pin and making it click into the right slot. It is also possible to adjust the distance between the arm cuffs and the handgrip: remove the nut and slide out the screw from the handgrip, insert the screw into the desired hole and fix by tightening the wing nut.

Standard crutches Cod. 27792

The height can be adjusted between 95 and 115 cm by pressing the pin and making it click into the right slot. It is also possible to adjust the distance between the arm cuffs and the handgrip: remove the nut and slide out the screw from the handgrip, insert the screw into the desired hole and fix by tightening the wing nut.

Standard crutches Cod. 27799

The height can be adjusted between 119 and 139 cm by pressing the pin and making it click into the right slot. It is also possible to adjust the distance between the arm cuffs and the handgrip: remove the nut and slide out the screw from the handgrip, insert the screw into the desired hole and fix by tightening the wing nut.

Standard crutches Cod. 27770

The height can be adjusted between 100 and 130 cm by pressing the pin and making it click into the right slot. It is also possible to adjust the distance between the arm cuffs and the handgrip: remove the nut and slide out the screw from the handgrip, insert the screw into the desired hole and fix by tightening the wing nut.

GIMA WARRANTY TERMS

The Gima 12-month standard B2B warranty applies.

	Medical Device compliant with Regulation (EU) 2017/745		Manufacturer	REF	Product code
	Caution: read instructions (warnings) carefully		Keep away from sunlight	LOT	Lot number
	Consult instructions for use		Keep in a cool, dry place	MD	Medical Device
	Authorized representative in the European community		Date of manufacture	UDI	Unique Identifier



Foshan Dongfang Medical Equipment Manufactory (LTD.)

5 Xingfu Road, Zone A, Shishan Industrial Park,
Nanhai District, Foshan City, China, 528000
Made in China



MedNet EC-Rep GmbH

Borkstrasse 10. 48163 Muenster.
Germany



FS937L - FS935S(M) - FS935S(S) - FS935S(L) - FS9252L

GIMA 27793 - 27798 - 27792
27799 - 27770

Importato da / Imported by / Importé par / Importiert von
Importado por / Importado por / Εισαγωγή από / مستورد عن طريق

Gima S.p.A. Via Marconi, 1 - 20060 Gessate (MI) Italy

gima@gimaitaly.com - export@gimaitaly.com

www.gimaitaly.com

