



GIMA

PROFESSIONAL MEDICAL PRODUCTS

ASPIRATORE CHIRURGICO TOBI
TOBI SUCTION ASPIRATOR
ASPIRATEUR TOBI
CHIRURGISCHER ABSAUGER TOBI
ASPIRADOR QUIRÚRGICO TOBI
APARELHO DE SUCÇÃO CIRÚRGICO TOBI

REF 28222



Gima S.p.A.
Via Marconi, 1 - 20060 Gessate (MI) Italy
gima@gimaitaly.com - export@gimaitaly.com
www.gimaitaly.com
Made in Italy



0476



TOBI SURGICAL ASPIRATOR is a medical device powered by 230V ~ 50Hz electricity, to be used for nasal, oral, tracheal aspiration of body fluids, such as mucus, phlegm and blood, in adults or children. Luminaire designed to offer ease of transport and continuous use.

Thanks to these features and its performance, this product is particularly suitable for use in hospital wards, for minor surgery applications and post-operative treatments at home. Built with a plastic body with high thermal and electrical insulation in compliance with European safety regulations, the appliance is supplied with a complete suction tank in sterilizable polycarbonate, with overflow valve, and is equipped with a suction regulator and vacuum gauge placed on the front panel.



GENERAL WARNING

Read instruction manual carefully before use.

The device is for use by qualified personnel (surgeon / professional nurse / assistant)

The use of the device at home is restricted to an adult in full possession of mental faculties and / or home carers

The instrument must not be disassembled. For technical service always contact Gima S.p.A.

IMPORTANT SAFETY RULES

1. Check the condition of the unit before each use. The surface of the unit should carefully be inspected for visual damage. **Check the mains cable and do not connect to power if damage is apparent;**
2. Before connecting the appliance always check that the electric data indicated on the data label and the type of plug used, correspond to those of the mains electricity to which it's to be connected;
3. Respect the safety regulations indicated for electrical appliances and particularly:
 - Use original components and accessories provided by the manufacturer to guarantee the highest efficiency and safety of the device;
 - The device can be used only with the bacteriological filter;
 - Never immerse the appliance into water;
 - Do not place or store the aspirator in places where it may fall or be pulled into the bathtub or washbasin. In the event it is accidentally dropped, do not attempt to remove the device from the water whilst the plug is still connected: disconnect the mains switch, remove the plug from the power supply and contact the GIMA technical service department. Do not attempt to make the device work before it has been thoroughly checked by qualified personnel and/or the GIMA technical service department.
 - Position the device on stable and flat surfaces in a way that the air inlets on the back aren't obstructed;
 - To avoid incidents, do not place the aspirator on unstable surfaces, which may cause it to accidentally fall and lead to a malfunction and/or breakage. Should there be signs of damage to the plastic parts, which may expose inner parts of the energised device, do not connect the plug to the electrical socket. Do not attempt to make the device work before it has been thoroughly checked by qualified personnel and/or the GIMA technical service department.
 - Don't use in the presence of inflammable substances such as anaesthetic, oxygen or nitrous oxide;
 - Don't touch the device with wet hands and always prevent the appliance coming into contact with liquids;
 - Don't leave the appliance connected to the power supply socket when not in use;
 - Don't pull the power supply cable to disconnect the plug remove the plug from the mains socket correctly;
 - Store and use the device in places protected against the weather and far from any sources of heat. After each use, it is recommended to store the device in its own box away from dust and sunlight.
 - In general, it is inadvisable to use single or multiple adapters and/or extensions. Should their use be necessary, you must use ones that are in compliance with safety regulations, however, taking care not to exceed the maximum power supply tolerated, which is indicated on the adapters and extensions.
 - Prevent children from using the device without proper supervision;
4. For repairs, exclusively contact technical service and request the use of original spare parts. Failure to comply with the above can jeopardise the safety of the device;
5. Use only for the purpose intended. Don't use for anything other than the use defined by the manufacturer. The manufacturer will not be responsible for damage due to improper use or connection to an electrical system not complying with current regulation.
6. The medical device requires special precautions regarding electromagnetic compatibility and must be installed and used in accordance with the information provided with the accompanying documents: the TOBI SUCTION UNIT device must be installed and used away from mobile and portable RF communication

devices (mobile phones, transceivers, etc.) that may interfere with the said device.

7. **Instrument and accessory discharging must be done according to current regulations in the country of use.**
8. **WARNING:** Do not change this equipment without the permission of the manufacturer GIMA S.p.A. None of electric or mechanical parts has been designed to be repaired by customers or end-users. Don't open the device, do not mishandle the electric / mechanical parts. Always contact technical assistance
9. Using the device in environmental conditions different than those indicated in this manual may harm seriously the safety and the technical characteristics of the same.
10. The medical device is in contact with the patient by means of a disposable probe (supplied with the device) furnished with the relative CE compliance certification according to the requirements of regulation ISO 10993-1: thus, no allergic reactions and skin irritations may occur.
11. The product and its parts are biocompatible in accordance with the requirements of regulation EN 60601-1.
12. Operation of the device is very simple and therefore no further explanations are required other than those indicated in the following user manual.
13. Use in Home-Care: Keep all accessories of the device out of reach of children under 36 months of age since they contain small parts that may be swallowed.
14. Do not leave the device unattended in places accessible to children and/or persons not in full possession of mental faculties as they may strangle themselves with the patient's tube and/or the power cable.



The manufacturer cannot be held liable for accidental or indirect damages should the device be modified, repaired without authorization or should any of its component be damaged due to accident or misuse. Any minimal modification / repair on the device voids the warranty and does not guarantee the compliance with the technical requirements provided by the MDD 93/42/EEC (and subsequent changes) and its normatives.

CONTRAINDICATIONS

- Before using the TOBI, consult the instructions for use: failure to read all the instructions in this manual can be harmful for the patient.
- The device cannot be used to drain chest fluids;
- The device must not be used for suction of explosive, corrosive or easily flammable liquids.
- TOBI is not suitable for MRI. Do not introduce the device in MRI environments.

TECHNICAL CHARACTERISTICS

Model	TOBI SUCTION ASPIRATOR
Typology (MDD 93/42/EEC)	Medical device Class IIa
Classification UNI EN ISO 10079-1	High Vacuum / Low Flow
Main Voltage	230 V ~ / 50 Hz
Power consumption	184 VA
Fuse	F 1 x 1.6A L 250 V
Maximum suction aspiration (without jar)	-75kPa (- 0.75bar) Regolable from -75kPa (-0.75 bar) to -10kPa (-0.10 bar)
Maximum flow (without jar)	16 l/min
Weight	2.5 Kg
Dimension	370 x 220 x 210 mm
Functioning	NON-STOP OPERATED
Accuracy of Vacuum Indicator	± 5%

Working Condition	Room temperature: 5 ÷ 35°C Room humidity percentage: 10 ÷ 93% RH Atmospheric pressure: 700 ÷ 1060 hPa
Conservation condition and Transport	Room temperature: -25 ÷ 70°C Room humidity percentage: 0 ÷ 93% RH Atmospheric pressure: 500 ÷ 1060 hPa

The technical specifications may change without notice.

CLEANING OF THE DEVICE

Use a soft dry cloth with not – abrasive and not – solvent detergents. To clean the device external parts always use a cotton cloth dampened with detergent. Don't use abrasive or solvent detergents. Before carrying out any cleaning and / or maintenance operation, disconnect the appliance from the power supply, unplugging it or turning off the switch on the device



PARTICULAR CARE SHOULD BE TAKEN TO ENSURE THAT THE INTERNAL PARTS OF THE EQUIPMENT DO NOT GET IN TOUCH WITH LIQUIDS. NEVER CLEAN THE EQUIPMENT WITH WATER.

During all clearing operations use protection gloves and apron (if need be, also wear a face mask and glasses) to avoid getting in contact with contaminating substances (after each utilization cycle of the machine).

ACCESSORIES SUPPLIES

- Complete aspiration jar 1000ml
- Conical fitting
- Tubes set 6mm x 10mm (transparent silicon)
- Aspiration probe CH20
- Antibacterial and hydrophobic filter

Available under request with different versions with complete jar 2000ml.

Anti-bacterial and hydrophobic filter: designed for the individual patient to protect patient and machine from cross-infections. Prevents the liquids, that come into contact with it, from passing through it. Replace it whenever you suspect that it may be contaminated and/or it becomes wet or discoloured. Replace the filter every time it is used if the suction pump is used on patients in unknown pathological situations and where an assessment of indirect contamination is not possible. The filter is not manufactured to be decontaminated, dismantled and/or sterilised. If, however, the patient's pathology is known and/or there is no risk of indirect contamination, the filter should be replaced after every work shift or once a month even if the device is not used.

Suction catheter: Single-use device to be used on a single patient. Do not wash or re-sterilize after use. Reuse may cause cross-infections. Don't use after lapse of the sell-by date

WARNING: Suction tubes for insertion in the human body purchased separately from the machine should comply with ISO 10993-1 standards on material biocompatibility.

Aspiration jar: The mechanical resistance of the component is guaranteed up to 30 cycles of cleaning and sterilization. Beyond this limit, the physical-chemical characteristics of the plastic material may show signs of decay. Therefore, we recommend that you to change it.

Silicone tubes: the number of cycles of sterilization and/or cleaning is strictly linked to the employment of the said tube. Therefore, after each cleaning cycle, it is up to the final user to verify whether the tube is suitable for reuse. The component must be replaced if there are visible signs of decay of the material constituting the said component.

Conical fitting: the number of cycles of sterilization and the number of cleaning cycles is strictly linked to the employment of the said component. Therefore, after each cleaning cycle, it is up to the final user to verify whether the fitting is suitable for reuse. The component must be replaced if there are visible signs of decay in the material constituting the said component.

Service life of the device: More than 850 hours of operation (or 3 years) in accordance with the standard conditions of testing and operation. Shelf life: maximum 5 years from the date of manufacture.

CLEANING OF ACCESSORIES

Before using the device, the manufacturer advises you to clean and/or sterilize the accessories. Washing and / or cleaning the autoclavable jar is to be carried out as follows:

- Wear protection gloves and apron (glasses and face mask if necessary) to avoid contact with contaminating substances.
- Disconnect the tank from the device and remove the said container from the support of the device.
- Separate all the parts of the cover (overflow device, washer).
- Disconnect all tubes from the jar and the protection filter.
- Empty and dispose of the contents of the suction vessel (also comply with regional regulations);
- Wash each part of the container from secretions under cold running water and then clean every single part in hot water (temperature not exceeding 60°C).
Once again, carefully wash each single part using, if necessary, a non-abrasive brush to remove any deposits.
- Rinse with hot running water and dry all parts with a soft cloth (non-abrasive).
- Dispose of the aspiration catheter according to the provisions of local laws and regulations.

The jug and lid can be further disinfected using a common disinfectant, strictly following the instructions and dilution values provided by the manufacturer. At the end of cleaning operations, leave to air dry in a clean environment.

The silicone aspiration tubes and the conical fitting may be carefully washed in hot water (temperature must not exceed 60°C). After cleaning, leave the parts to dry in an open, clean environment.

When cleaning is complete, reassemble the container for liquid aspirations according to the following procedure:

- Place the overflow valve into its seat in the cover (under VACUUM connector)
- Insert floating valve keeping the o-ring towards the opening of the cage
- Place the o-ring into its seat around the cover
- After completing assembling operations always make sure that cover seals perfectly to avoid vacuum leakages or liquid exit

The jar and the cover can be autoclaved by placing the parts into the autoclave and running one sterilization stem cycle at 121°C (1 bar relative pressure – 15 min) making sure that the jar is positioned upsidedown. Mechanical resistance of the jar is guaranteed up to 30 cycles of sterilization and cleaning at the indicated conditions (EN ISO 10079-1). Beyond this limit the physical-mechanical characteristics of the plastic may decrease and replacement of the part is therefore recommended.

After sterilization and cooling at environment temperature of the parts make sure that these are not damaged. The aspiration tubes can be sterilized on autoclave using a sterilization cycle at 121°C (1 bar relative pressure – 15 min).

The conical connector can be sterilized on autoclave using a sterilization cycle at 121°C (1 bar relative pressure – 15 min).



DO NOT WASH, STERILIZE OR PUT IN AUTOCLAVE THE ANTIBACTERIAL FILTER

PERIODICAL MAINTENANCE CHECKS

The TOBI SUCTION ASPIRATOR does not need maintenance or lubrication.

It is, however, necessary to inspect the unit before each use. With regard to training, given the information contained in the user manual and since it is easy to understand the said device, it doesn't appear to be necessary.

Unpack the instrument and always check integrity of plastic parts and feeding cable, they might have been damaged during previous use. Connect the cable to electrical network and turn the switch on.

Close the aspirator outlet with your finger and with suction regulator at maximum check that the vacuum indicators reaches at least -75kPa (-0.75 bar). Rotate the knob from right to left. The vacuum indicator should go down -25kPa (-0.25 bar).

Check that no loud noises are present. A protection fuses (**F 1 x 1.6 A L 250V**) reachable from exterior and situated in the plug protects the instrument. For fuses replacing, always the type and the range.

Before changing the fuse, disconnect the plug from the power supply socket.

Fault type	Cause	Solution
1. The suction unit doesn't work	Cable is damaged External power source failure	Replace the cable Check the external power source
2. No aspiration	Jar Cap not screwed on correctly	Unscrew the cap, and re-screw it
3. No aspiration	Lid seal not in its seat	Unscrew the cap and insert the seal properly in its seat
4. The Vacuum power on the patient side is either very low or absent	<ul style="list-style-type: none"> • Vacuum regulator set to minimum • Protection filter blocked or damaged • Connection tubes blocked, kinked or disconnected • Shut-off valve blocked or damaged • Pump motor damaged 	<ul style="list-style-type: none"> • Turn the vacuum regulator clockwise and check the value of the vacuum on the gauge • Replace the filter • Replace or reconnect the tubes, check the jar connections • Empty the jar, or disconnect the tube from the jar and unblock the shut-off valve. The unit will only work in the upright position • Refer to authorised service personnel
5. The float doesn't close	If the cap has been washed, ensure that the float is not partially detached	Fit the float into its place
6. The float doesn't close	The float it's covered by dirty material	Unscrewed the cap, leave the and put in on autoclave
7. Low suction	Foam inside the jar	Fill the jar to 1/3 full of ordinary water
8. No aspiration due to flow leakage of mucus	Filter blocked	Replace filter
Faults 1 - 2 - 3 - 4 - 5 - 6 - 7 - 8	None of the procedures have achieved the desired results	Contact GIMA customer service

If the overflow security system it's activated, don't proceede with the liquid aspiration.

If the overflow security system doesn't work there are two cases:

1° case – If the overflow security system doesn't work the aspiration will be stopped by the bacteriological filter who avoid the liquid penetration inside the device.

2° case – If both the security system doesn't work, there is the possibility that liquid comes inside the device, in this case return the device to GIMA technical service.

Gima S.p.A. will provide upon request electric diagrams, components list, descriptions, setting instructions and any other information that can help the technical assistance staff for product repair.



BEFORE EVERY CHECKING OPERATION, IN CASE OF ANOMALIES OR BAD FUNCTIONING, PLEASE CONTACT GIMA TECHNICAL SERVICE. GIMA DOES NOT GIVE GUARANTEE IF INSTRUMENT, AFTER THE TECHNICAL SERVICE CHECKING, APPEARS TO BE TAMPERED.


INSTRUCTION FOR USE

- The device must be checked before each use in order to detect malfunctions and / or damage caused by transport and / or storage.
- The working position must be such as to allow one to reach the control panel and to have a good view of the empty indicator, the jar and the antibacterial filter.
- It is recommended not to keep the device in your hands and / or to avoid prolonged contact with the body of apparatus.

WARNING: For proper use, place the aspirator on a flat, stable surface in order to have the full volume of use of the jar and better efficiency of the overflow device.

- Connect one end of the short silicon tube, with antibacterial filter, to the suction connector on the lid of the jar.
- The other tube, connected to the filter on one end, must be attached to the connector on the flask cover to which the float is secured inside. (overflow device). The overflow device starts working (the float closes the connector on the cover) when the maximum volume of liquid is reached, so no liquid can enter the machine (90% of the flask's total volume), thus ensuring that the liquid cannot penetrate inside the machine. The device must be used on a flat work top.

Filter assembling



Make sure the filter is assembled with the arrows on the side of the patient.

WARNING: The inside of the medical device must be regularly checked for the presence of liquids or other visible contamination (secretions). In the presence of liquids or other visible contamination, immediately replace the medical device due to the risk of an insufficient vacuum flow rate.

These products have been designed, tested and manufactured exclusively for single patient use and for a period no longer than 24 hours.

- Connect the long silicone tube to the "PATIENT" jar outlet
- Connect the other end of the long silicon tube to the probe plastic connector, then connect the suction probe to it.
- Connect the power cord to the device then connect the plug to the electrical mains supply.
- Push switch on position I to start suction.
- Unscrew the lid of the jar and fill the jar 1/3 full of ordinary water (this assists the unit to reach peak vacuum performance and makes clean-up easier) then re-screw the lid on the jar correctly.
- During operation the jar has to be in vertical position to avoid overflow valve to cut off aspiration. Should this happen, switch off the device and disconnect the tube from the jar cover (from "VACUUM" outlet).
- Once finished push switch on O position and unplug.
- Remove the accessories and clean.
- At the end of each use, place the device in its box away from dust.

WARNING: The power supply cable plug is the element of separation from the electrical mains system: even if the units equipped with a special on / off switch button, the power supply plug must be kept accessible once the device is in use so as to allow a further method of disconnection from the mains supply system.



**NEVER USE THE DEVICE WITHOUT JAR AND / OR PROTECTION FILTER
MAKE SURE THAT CHILDREN AND/OR MENTALLY ILL PEOPLE DO NOT USE THE DEVICE WITHOUT
ADULT SURVEILLANCE**

RISK OF ELECTROMAGNETIC INTERFERENCE AND POSSIBLE REMEDIES


This section contains information regarding the conformity of the compliance with the EN 60601-1-2 Standard (2014).

The TOBI surgical aspirator is an electro-medical device that requires particular precautions regarding electro-magnetic compatibility and which must be installed and commissioned according to the electro-magnetic compatibility information supplied. Portable and mobile radio communication devices (mobile phones, transceivers, etc.) may interfere with the medical device and should not be used in close proximity with, adjacent to or on top of the medical device. If such use is necessary and unavoidable, special precautions should be taken so that the electro-medical device functions properly in its intended operating configuration (for example, constantly and visually checking for the absence of anomalies or malfunctions). The use of accessories, transducers and cables differing from those specified, with the exception of transducers and cables sold by the appliance and system manufacturer as spare parts, can lead to an increase in emissions or in a decrease of the immunity of the device or system. The following tables supply information regarding the EMC (Electromagnetic Compatibility) characteristics of the electro-medical device.

Guidance and manufacturer's declaration – electromagnetic Emissions		
The TOBI SUCTION UNIT is intended for use in the electromagnetic environment specified below. The customers or the user of the TOBI SUCTION UNIT should make sure that it's used in such an environment.		
Emissions test	Compliance	Electromagnetic environment - guidance
Irradiated / Conducted emissions CISPR11	Group 1	The TOBI SUCTION UNIT only used RF energy only for its internal functioning. Therefore, its RF emissions are very low and are not cause interference in proximity of any Electronic appliances.
Irradiated / Conducted emissions CISPR11	Class [B]	The TOBI SUCTION UNIT can be used in all environments, including domestic and those connected directly to the public mains distribution that supplies power to environments used for domestic scopes.
Harmonic emissions EN 61000-3-2	Class [A]	
Voltage fluctuations / flicker emissions EN 61000-3-3	Complies	

Guidance and manufacturer's declaration – Immunity Emissions			
The TOBI SUCTION UNIT is intended for use in the electromagnetic environment specified below. The customers or the user of the TOBI SUCTION UNIT should make sure that it's used in such an environment.			
Immunity Test	Level indicated by the EN 60601-1-2	Compliance Level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) EN 61000-4-2	+/-8kV on contact +/-15kV in air	The device doesn't change its state	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient / burst EN 61000-4-4	± 2kV power supply lines ± 1kV for input / output lines	The device doesn't change its state	Mains power quality should be that of a typical commercial environment or hospital.
Surge EN 61000-4-5	± 1kV differential mode +/-2 kV ordinary mode	The device doesn't change its state	Mains power quality should be that of a typical commercial environment or hospital.
Loss of voltage, brief voltage interruptions and variations EN 61000-4-11	<5% U _T (>95% dip U _T) for 0,5 cycle 40% U _T (60% dip U _T) for 5 cycle 70% U _T (30% dip U _T) for 25 cycle <5% U _T (>95% dip U _T) for 5 sec	- -	Mains power quality should be that of a typical commercial environment or hospital. If the user of the TOBI SUCTION UNIT request that the appliance operates continuously, the use of a continuity unit is recommended.
Magnetic field with network frequency (50/60 HZ) EN 61000-4-8	30A/m	The device doesn't change its state	The power frequency magnetic field should be measured in the intended installation location to make sure that it's sufficiently low.
Note U _T is the value of the power supply voltage			

Guidance and manufacturer's declaration – Immunity Emissions			
The TOBI SUCTION UNIT is intended for use in the electromagnetic environment specified below. The customers or the user of the TOBI SUCTION UNIT should make sure that it's used in such an environment.			
Immunity Test	Level indicated by the EN 60601-1-2	Compliance Level	Electromagnetic environment - guidance
Conducted Immunity EN 61000-4-6	3Vrms 150kHz to 80Mhz (for non life-supporting devices)	V1 = 3 V rms	The portable and mobile RF communication devices, including cables, must not be used closer to the TOBI SUCTION UNIT device, than the separation distance calculated by the equation applicable to the transmitter frequency. Recommended separation distance $d = \left[\frac{3,5}{V^1} \right] \sqrt{P}$ $d = \left[\frac{12}{E^1} \right] \sqrt{P} \quad \text{from 80 MHz to 800MHz}$ $d = \left[\frac{23}{E^1} \right] \sqrt{P} \quad \text{from 800 MHz to 2.7 GHz}$
Radiated Immunity EN 61000-4-3	10V/m 80MHz to 2.7GHz (for non life-supporting devices)	E1 = 10 V / m	

		<p>Where P is the maximum nominal output voltage of the transmitter in Watt (W) depending on the manufacturer of the transmitter and the recommended separation distance in metres (m). The intensity of the field from the fixed RF transmitters, as determined by an electro-magnetic study of the site^{a)}, could be lower than the level of conformity of each frequency interval^{b)}. It is possible to check for interference in proximity to devices identified by the following symbol:</p> 
--	--	---

Note 1: At 80 MHz and 800 MHz the interval with the highest frequency is applied

Note 2: These guide lines may not be applicable in all situations. The electro-magnetic propagation is influenced by the absorption and by reflection from buildings, objects and people.

a) The field intensity for fixed transmitters such as the base stations for radiotelephones (mobile and cordless) and terrestrial mobile radio, amateur radio devices, radio AM and FM transmitters and TV transmitters can not be theoretically and accurately foreseen. To establish an electro-magnetic environment generated by fixed RF transmitters, an electro-magnetic study of the site should be considered. If the field intensity measured in the place where the device will be used surpasses the above mentioned applicable level of conformity, the normal functioning of the device should be monitored. If abnormal performance arises, additional measures such as changing the device's direction or positioning may be necessary.

b) The field intensity on an interval frequency of 150 kHz to 80 MHz should be less than 3 V/m.

Recommended separation distance between portable and mobile radio-communication devices and the monitor

The TOBI SUCTION UNIT surgical aspirator is intended to operate in an electro-magnetic environment where RF irradiated interferences are under control. The client or operator of the TOBI SUCTION UNIT device can help prevent electro-magnetic interference by keeping a minimum distance between the portable and mobile RF communication devices (transmitters) and the TOBI SUCTION UNIT device, as recommended below, in relation to the radio-communication maximum output power.























Maximum nominal output power of the Transmitter W	Separation distance from the frequency transmitter (m)		
	150KHz to 80MHz $d = \left[\frac{3,5}{V^1} \right] \sqrt{P}$	80MHz to 800MHz $d = \left[\frac{12}{E^1} \right] \sqrt{P}$	800MHz to 2,7GHz $d = \left[\frac{23}{E^1} \right] \sqrt{P}$
0,01	0,12	0,12	0,23
0,1	0,38	0,38	0,73
1	1,2	1,2	2,3
10	3,8	3,8	7,3
100	12	12	23

For transmitters with a maximum nominal output power not shown above, the recommended separation distance in metres (m) can be calculated using the equation applicable to the transmitter frequency, where P is the maximum nominal output power of the transmitter in Watt (W) depending on the transmitter's manufacturer.

Note 1: At 80 MHz and 800 MHz the interval with the highest frequency is applied

Note 2: These guide lines may not be applicable in all situations. The electro-magnetic propagation is influenced by the absorption and by the reflection from buildings, objects and people

SYMBOLS

	Caution: read instructions (warnings) carefully		Follow instructions for use
	Keep in a cool, dry place		Keep away from sunlight
	Manufacturer		Date of manufacture
	Product code		Lot number
	Medical Device complies with Directive 93/42/EEC		Type BF applied part
	WEEE disposal		Class II applied
	Serial number		Temperature limit
	Fuse		Atmospheric Pressure limit
	Covering Protection rate		Alternating current
	ON		Mains frequency
	OFF		Humidity limit



Disposal: The product must not be disposed of along with other domestic waste. The users must dispose of this equipment by bringing it to a specific recycling point for electric and electronic equipment.

GIMA WARRANTY TERMS

The Gima 12-month standard B2B warranty applies.

