

FIG.-IM.-BILD-GRAFIC-RYS. A



The product is supplied for the foot control of compatible electronic units.

Verify the status of the product before the use. Throw away the devective products.

This product is compatible with versions of equipment that require it, refer to the instructions for the unit.

The connections must be executed by trained staff.

Installation:

- Insert the product connector on specific electronic unit connector, paying attention to the direction of insertion (fig. A -part. 1-2).
- -to lock the product, turn the dial on the unit (fig. A part. 3).

Turn on the unit and perform functional tests by acting on the pedal (fig. A part 4).







