

OPERATION AND MAINTENANCE MANUAL

E/N SERIES TRIS-LED

MINOR SURGICAL LUMINAIRE (TREATMENT LUMINAIRE)

Introduction

Please read this manual carefully before using the Product, so as to protect **"the Technical Service Personnel"** and **"the Operator"** from any injury.

Marking

This appliance is a Class I medical device pursuant to REGULATION (EU) 2017/745 on medical devices (Annex VIII) as amended and integrated.

Compliance

The manufacturer declares that this Product complies with Annex I (General Safety and Performance Requirements) of REGULATION (EU) 2017/745 as amended and integrated and certifies such conformity by affixing the CE marking.

Validity of manual

This installation manual is valid for the following models:

- PENTALED 30E in ceiling, wall and mobile versions;
- PENTALED 81 in ceiling version;
- PENTALED 30N in ceiling, wall and mobile versions;
- PENTALED 63N in ceiling version;
- TRIS-LED in ceiling, wall and mobile versions.

Customer Service

The customer service is at your disposal in case of Product details, information concerning its use, identification of spare parts being required and for any other queries you might have concerning the appliance, for ordering spares and for matters relating to assistance and warranty.

- RIMSA P. LONGONI SRL
- Via Monterosa 18
- I-20831 Seregno MB
- Tel.: ++39 0362 325.709
- Fax: ++39 0362 328.559
- E-mail: info@rimsa.it

Copyright

The reproduction and translation, including partial, of any part of this manual is forbidden without the written permission of RIMSA.

Translations

The original language of this manual is ITALIAN. For all translations, reference must be made to the original manual language.

Index of contents

KEY.....	5
1 GENERAL SAFETY INFORMATION	6
2 Importance of personal safety	6
2.1 Intended use	6
2.2 Safety conditions (secondary effects)	7
2.3 Environmental conditions	7
3 General information	8
3.1 Operator qualifications	8
3.2 Patient population and body interactions	8
3.3 Graphic symbols used in this operation and maintenance manual	9
3.4 Graphic symbols used on the Product	9
4 Precautions for the Product operator	10
4.1 Personnel awareness obligation	10
4.2 Warranty and liabilities	10
5 Product description and operation	11
5.1 Product description E/N SERIES	11
5.2 Product description E/N SERIES	13
5.3 Description of operation	15
5.3.1 E SERIES control keyboard	15
5.3.2 N SERIES control keyboard	16
5.3.3 TRIS-LED control keyboard	16
5.4 Product handling	17
5.4.1 Brakes for mobile version	21
5.4.2 Moving the stand	22
5.5 Checks to be made every time before use	22
6 Cleaning and disinfecting	22
6.1 Application method	22
6.2 Cleaning the Product	23
6.3 Product disinfecting	23
6.4 Handpiece sterilization	24
7 Adjustment and maintenance	25
7.1 Swinging arm adjustment	25
7.2 Clutches adjustment	26
7.2.1 E/N SERIES clutches adjustment	26
7.2.2 TRIS-LED clutches adjustment	26
7.3 Periodical checks to be performed on the Product	27
7.4 Routine maintenance	27
7.5 Repairs	29
7.6 Disposal after use	30
7.7 Spare parts list	31
8 Technical properties	32
8.1 E SERIES technical properties	32
8.2 N SERIES technical properties	34
8.3 TRIS-LED technical properties	36
9 Declaration of Conformity of the Manufacturer	38
10 EMC Declaration	40
11 Warranty Certificate	45

BG	За да поискате наръчника на този език, изпратете имейл на адрес info@rimsa.it .
CS	Chcete-li si vyžádat příručku v tomto jazyce, zašlete e-mail na adresu info@rimsa.it .
DA	Hvis du ønsker at få manualen på dette sprog, bedes du sende en e-mail til info@rimsa.it .
DE	Um das Handbuch in dieser Sprache anzufordern, senden Sie bitte eine E-Mail an info@rimsa.it .
EL	Για να ζητήσετε το εγχειρίδιο σε αυτή τη γλώσσα, στείλτε μήνυμα ηλεκτρονικού ταχυδρομείου στη διεύθυνση info@rimsa.it .
ES	Para solicitar el manual en este idioma, envíe un correo electrónico a info@rimsa.it .
ET	Selles keeles käsiraamatu tellimiseks saatke palun e-kiri aadressile info@rimsa.it .
FI	Jos haluat käsikirjan tällä kielellä, lähetä sähköpostia osoitteeseen info@rimsa.it .
FR	Pour demander le manuel dans cette langue, veuillez envoyer un e-mail à info@rimsa.it .
GA	Chun an lámhleabhar sa teanga seo a iarraidh, seol r-phost chuig info@rimsa.it .
HR	Da biste zatražili priručnik na ovom jeziku, pošaljite e-mail na info@rimsa.it .
HU	A kézikönyv ezen a nyelven történő igényléséhez kérjük, küldjön e-mailt a info@rimsa.it címre.
LT	Norėdami prašyti vadovo šia kalba, siųskite el. laišką adresu info@rimsa.it .
LV	Lai pieprasītu rokasgrāmatu šajā valodā, lūdzu, sūtiet e-pastu uz adresi info@rimsa.it .
MT	Biex titlob il-manwal f'din il-lingwa, jekk jogħġbok ibgħat e-mail lil info@rimsa.it .
NL	Om de handleiding in deze taal aan te vragen, kunt u een e-mail sturen naar info@rimsa.it .
PL	Aby zamówić podręcznik w tym języku, należy wysłać wiadomość e-mail na adres info@rimsa.it .
PT	Para solicitar o manual nesta língua, envie por favor um e-mail para info@rimsa.it .
RO	Pentru a solicita manualul în această limbă, vă rugăm să trimiteți un e-mail la info@rimsa.it .
SK	Ak chcete požiadať o príručku v tomto jazyku, pošlite e-mail na adresu info@rimsa.it .
SL	Če želite zahtevati priročnik v tem jeziku, pošljite e-pošto na naslov info@rimsa.it .
SV	Om du vill ha handboken på detta språk skickar du ett e-postmeddelande till info@rimsa.it .

PRODUCT**KEY**

ME (Medical Electrical) EQUIPMENT to which this manual refers is a **MINOR SURGICAL LUMINAIRE (TREATMENT LUMINAIRE)**. For ease of description, in this manual this ME EQUIPMENT will be called "**Product**".

OPERATOR

Professional medical personnel (e.g., professional health personnel, expert person assisting the patient).

**RESPONSIBLE
ORGANIZATION**

Entity accountable for the use and maintenance of an ME equipment or ME system (e.g., a hospital, an individual doctor or a non-expert person). Preparation and awareness are included in use.

**TECHNICAL
SERVICE
PERSONNEL**

The personnel (individuals or entity accountable to the responsible organization) that installs, assembles, maintains or repairs the equipment. Under certain circumstances, the safety of such persons depends on their knowledge and awareness and ability to take appropriate precautions when gaining access to hazardous parts partially. By way of example only, the following professional figures are deemed as SERVICE PERSONNEL:

- ⇒ Construction Engineer, Draughtsman, Building firm duly registered in the professional Register (for the masonry works)
- ⇒ Electrical Engineer Electro-technical expert qualified to work as an electrician (for the electrical works)

1 GENERAL SAFETY INFORMATION

This manual is an integral part of the Product as indicated by REGULATION (EU) 2017/745 and subsequent amendments and supplements. Read and keep this Operation and Maintenance Manual close to the Product.

RIMSA disclaims all liability for any injury to persons or damage to property caused by the USE or MAINTENANCE of the Product by persons who are not OPERATORS or TECHNICAL SERVICE PERSONNEL. The Product is an ME Medical Electrical equipment and therefore falls within the field of application of the IEC 62353 standard.

To avoid any risk of electric shocks, the Product must only be connected to mains supplies with earth protection.



Electric shock risk.

2 Importance of personal safety

2.1 Intended use

MINOR SURGICAL LUMINAIRE (TREATMENT LUMINAIRE)

The Product is a medical device designed for use in operating theatres within the PATIENT AREA, with short-term duration, active, non invasive, designed to locally light up the patient's body for treatments and diagnosis which can be interrupted without any HAZARD for the PATIENT in case of failure of the light.

A combination of two or more surgical lamps used in the operating theatre and required for treatment and diagnosis makes up a SURGICAL LUMINAIRE SYSTEM.

The Product correctly lights up the operating range from a distance of about 70 – 140 cm from the patient area.

In the event of overlapping lamps, a temperature increase would ensue in the patient area with consequent risk of dehydration and tissue damage.

In case of a reduction in blood flow with start of tissue dehydration, reduce light intensity.

Operating field

Undesired effects of overlapping light fields



Possibility of tissue dehydration and damage.

Optical safety



Electromagnetic disturbance

Incorrect use



Improper use of mobile version



2.2 Safety conditions (secondary effects)

- Do not direct the light source into the patient's and/or operator's eyes.
- When Product use is restricted to the face (maxilla-facial surgery, plastic surgery, ear-nose-throat surgery) the patient's eyes must be covered with adequate protection.
Failure to follow such precautions could cause glare and potential damage to the retina.

To avoid any significant risk of reciprocal interference due to the presence of the Product during specific exams or treatments, refer to section 10 of the Manual.

- Never place and/or hang anything on the Product.
Failure to follow such precaution could result in such objects falling in the operating area.
- Never hang on the Product with the body weight of a person.
Failure to follow such precaution could damage the Product structure.
- Never cover the head of the Product during operation to prevent overheating.
- Avoid the Product parts colliding with one another or other nearby equipment.

Knocks could cause the detachment of plastic parts or paint from the Product which could fall in the patient area.

In the case of the mobile version, do not rest, push or lie on the product. Failure to comply could result in damage to the product and to devices nearby and injury to staff members.

2.3 Environmental conditions

- The Product is not suitable for use in explosion-risk areas.
- The Product is not suitable for use wherever there are inflammable mixes of anaesthetics with air, oxygen or N₂O (laughing gas).
- The Product is not suitable for use in environments rich in oxygen and use is not intended in the presence of inflammable agents.
- During operation, the ambient temperature must be between 10°C and 40°C.
- Relative humidity must be between 30% and 75%.
- Atmospheric pressure must be between 700 and 1060hPa.

3 General information

3.1 Operator qualifications

Qualification of personnel in charge of operating on the Product:

Professional medical personnel.

Properly trained medical and paramedical personnel.

Qualified technician with required technical-professional skills.

RIMSA or technical service personnel, the latter only for the fuse change.

RIMSA or authorized Dealer.

Comply with applicable laws on waste disposal. This product must not be disposed of in standard waste disposal bins. To avoid risks for the environment and health deriving from the dispersion of polluting substances in the environment, separate the various internal component parts such as iron, aluminium, plastic and electrical material, and dispose of these through authorized channels so as to ensure correct recycling.

3.2 Patient population and body interactions

The intended use makes the Product suitable for all types of population without constraints of age, weight, health or medical conditions. Patients can be awake or unconscious, in local or total anaesthesia. Patient population can also be made of animals.

An active patient could only accidentally touch the head and the swinging arm of the device, while this is not possible in case of unconscious or disable patients.

The operator touches the device necessary on the sterilisable handpiece and function control keyboard, and occasionally on the enclosure. In any case, he/she probably wears personal protective equipment (PPE) so it isn't a direct contact.

Use
Cleaning
Routine maintenance
Special maintenance

Assistance
Demolition

Patient population

Patient interaction

Operator interaction

3.3 Graphic symbols used in this operation and maintenance manual

The following safety measures must be put in place during Product installation, use and servicing.

To emphasize their importance, a number of safety precautions are repeated throughout the manual.

Follow the safety precautions before using or repairing the Product. Carefully abiding by the safety precautions improves the ability to use the Product safely and correctly and helps prevent incorrect maintenance which could be hazardous and cause damage. The safety measures are approximate and not exhaustive; the Operator, the Responsible Organization and the Technical Service Personnel must develop their capacities to upgrade and integrate them.



General warning signal



General mandatory code of conduct signal



General prohibition signal

3.4 Graphic symbols used on the Product

Below are the symbols to be found on the Product:

CE marking indicating the Product complies with REGULATION (EU) 2017/745 and subsequent amendments and supplements



Date of manufacture (month and year)



Manufacturer's address



Fuses used in the device



Comply with the instructions for use



Medical Device



Reference number



Serial number



Disposal




'N'

'L'

'I'

'O'



Protection earth

Neutral lead connection point

Line lead connection point

ON

OFF

Standby and switch-on

Pushing, resting on or lying on the product is forbidden

No stepping on surface

Only move the product after lowering the arm

4 Precautions for the Product operator

4.1 Personnel awareness obligation

The Responsible Organization must instruct the Operator on how to use, clean and service the Product.

The instructions must be provided in written form on the basis of this Manual.

4.2 Warranty and liabilities

RIMSA disclaims all liability as regards unreliable Product operation in the following cases:

- The Product has not been used for its intended purpose and in conformity with the operating instructions.
- Authorized modifications and repairs have not been performed by TECHNICAL SERVICE PERSONNEL.

Operator Instructions

5 Product description and operation

5.1 Product description E/N SERIES


The Product is available in various versions:


- **ceiling version**
- **double ceiling version**
- **wall version (only PENTALED 30E/30N)**
- **mobile version (only PENTALED 30E/30N)**

SINGLE CEILING version: ceiling cover (1), ceiling anchor tube (2), power supply (3), horizontal arm (4), swinging arm (5), fork (6), lamp head (7), control keyboard (8), sterilisable handpiece (9).


DOUBLE CEILING version: ceiling cover (1), ceiling anchor tube (2), power supply (3), horizontal arm (4), swinging arm (5), fork (6), lamp head (7), control keyboard (8), sterilisable handpiece (9).

Versions





WALL version for 30E, 30N lamps: wall box (1), power plug (2), horizontal arm (3), swinging arm (4), lamp head (5), control keyboard (6), sterilisable handpiece (7).



MOBILE version for 30E, 30N lamps: base with wheels (1), power plug (2), inferior stem (3), superior stem (4), swinging arm (5), lamp head (6), function control keyboard (7), sterilisable handpiece (8).

Separable parts

Sterilisable handpiece Refer to Section 6.4 for assembly/disassembly instructions.

Versions


5.2 Product description TRIS-LED


The Product is available in various versions:

- **ceiling version**
- **double ceiling version**
- **wall version**
- **mobile version**


SINGLE CEILING version: ceiling cover (1), ceiling anchor tube (2), power supply (3), horizontal arm (4), swinging arm (5), lamp head (6), control keyboard (7), sterilisable grip (8).

DOUBLE CEILING version: ceiling cover (1), ceiling anchor tube (2), power supply (3), horizontal arm (4), swinging arm (5), lamp head (6), control keyboard (7), sterilisable grip (8).





WALL version: wall box (1), power plug (2), horizontal arm (3), swinging arm (4), lamp head (5), control keyboard (6), sterilisable grip (7).



MOBILE version: base with wheels (1), power plug (2), inferior stem (3), superior stem (4), swinging arm (5), lamp head (6), function control keyboard (7), sterilisable grip (8).


Separable parts

Sterilisable handpiece Refer to Section 6.4 for assembly/disassembly instructions.

Main switch

CAUTION

E SERIES control keyboard



Lighted area

5.3 Description of operation

Wall and mobile version lamps feature a green light switch for general switching on and off.

In case of single and double ceiling versions position the thermal magnetic switch near the Product so that it can be switched off in case of need.

In case of wall and mobile versions do not position the device so it is hard to reach and remove the power plug in case of an emergency.


5.3.1 E SERIES control keyboard

The membrane keyboard is located on the yoke. It can control:

- lamp switch-on and switch-off by means of the "I/O" key (1);
- adjustment of light intensity by dragging your finger over the bar or touching the sun symbol keys (2). The level of intensity achieved is indicated by means of 5 green microleds;
- selection of colour temperature between 2 values: 4500K and 5000K by pressing the keys with the letter K indicating the value (3);
- enabling the "Endoled" function, using the key with the letter E (4). This function is only available with lamp off;
- adjustment of light range (increase-decrease) by means of key (5) which increases or reduces the range.


The Product has been designed to be able to regulate the light diameter electronically by means of the key (5) provided.

N SERIES control keyboard



Light field adjustment

TRIS-LED control keyboard



Light field adjustment

5.3.2 N SERIES control keyboard

The membrane keyboard is located on the yoke. It can control:

- lamp switch-on and switch-off by means of the "I/O" key (1);
- adjustment of light intensity by dragging your finger over the bar or touching the sun symbol keys (2). The level of intensity achieved is indicated by means of 5 green microleds;
- selection of colour temperature between 2 values: 4500K and 5000K by pressing the keys with the letter K indicating the value (3);
- enabling the "Endoled" function, using the key with the letter E (4). This function is only available with lamp off.

The mechanical adjustment of the light field can be performed by turning the sterilisable handpiece. Turn clockwise to widen the light field in the lighted area (patient area) and anticlockwise to narrow it.

5.3.3 TRIS-LED control keyboard


The function control keyboard is mounted on the Product cupola. It can control:

- lamp switch-on and switch-off by means of the "I/O" key (1);
- adjustment of light intensity by dragging your finger over the bar or touching the sun symbol keys (2). The level of intensity achieved is indicated by means of 5 green microleds (3);


Through the central sterilisable handle is possible to adjust the light field. Rotating clockwise or counter clockwise the central handle is possible to enlarge or reduce the light field diameter (patient area).

5.4 Product handling





SINGLE ceiling model PENTALED 30E/30N






SINGLE ceiling model PENTALED 81/63N




SINGLE ceiling model TRIS-LED







DOUBLE lamp model





Wall model PENTALED 30E/30N





Wall model TRIS-LED



Mobile model PENTALED 30E/30N



Mobile model TRIS-LED



The Product can be moved using the sterilisable handpiece.


For PENTALED models the Product can also be moved using the the side handles.

By pressing the keys on the keyboard, the previously described control functions can be enabled.

For N SERIES models, it's possible to rotate clockwise or counterclockwise the sterilisable handpiece to adjust the diameter of the light field and the focus.


Also for TRIS-LED models, it's possible to rotate clockwise or counterclockwise the sterilisable handpiece to adjust the light field and the focus.





RECOMMENDED WORK DISTANCE

To optimize light intensity, the product is best used at a distance of 1m.



The Product nevertheless also ensures a good light intensity at a distance between 70cm and 140cm.



Risk of damaging pedal.



5.4.1 Brakes for mobile version

The mobile version has 4 wheels with pedal brake. This are used to block the Product in the required position.

Press the brake pedal with your foot, without applying too much pressure.

Do not kick the brake pedal and do not press continuously once the stop position has been reached.

To disengage the brake, lift the pedal with your foot.



Risk of lamp overturning.

5.4.2 Moving the stand

Whenever the stand has to be moved, make sure the swinging arm is moved downwards.

Failure to do so could cause the lamp to overturn.

5.5 Checks to be made every time before use

To make sure the Product is safe and provides a correct diagnosis, every time before use, the operator must check:

- The lamp has been correctly disinfected;
- The emitted light is stable and of adequate intensity;
- The swinging arm maintains correctly its position;
- The cupola maintains correctly its position.

6 Cleaning and disinfecting

The responsible organization must comply with the rules (standards and directives) concerning hygiene, disinfection and sterilization laid down by the relevant national commission.

6.1 Application method


Before proceeding to clean / disinfect the Product, make sure it is off and cannot be switched back on.

Allow the lamp to cool down and only clean it when it is cold.


Protect the Product from water spray and detergents and do not clean it in direct contact with liquids.

Do not spray detergent / disinfectant directly on Product.

Application method



Spray the detergent / disinfectant on a cloth so as to dampen it.



Afterwards wipe the Product with the cloth.

Failure to comply with the above instructions could cause:

- detaching of paint with possible accidental dropping of such paint into the patient area;
- early ageing of the plastic parts with consequent weakening and the possibility of breakages;
- tarnishing of the protection screens and glass.

6.2 Cleaning the Product

We recommend you to clean the Product every day.

- Do not use sharp, pointed or abrasive objects, to avoid the risk of damaging surfaces.
- Do not pour liquids directly on the Product.
- Clean the Product with a damp, but not wet, cloth.
- Clean with suitable detergents with low alkaline content and chlorine free. Do not use abrasive products, petrol, paint thinners, alkaline detergents, acids, containing alcohol or aldehydes.
- Dose the detergents strictly according to the percentage indications shown on the manufacturer's technical sheet, being careful that no liquids penetrate into the joints of the various Product parts, with special care give to the reflector and supporting structure.

6.3 Product disinfecting

We recommend you to disinfect the Product every time before use. Disinfectants can contain substances that are harmful for the health; use disinfectants indicated by the national commission for hygiene and disinfection, according to the hygienic standards adopted by the Responsible Organization.

- Do not use sharp, pointed or abrasive objects, to avoid the risk of damaging surfaces.
- Do not pour liquids directly on the Product.
- Disinfect the Product with a damp but not wet cloth.
- Use appropriate disinfectants with low alcohol content.
- To prevent damaging the stainless-steel and aluminium parts, use only disinfectants that do not contain chlorine or halogens.
- Dilute the disinfectants in strict accordance with the percentage indications on the manufacturer's technical data sheet, being careful no liquids penetrate into the joints of the various parts of the Product, with special attention for the reflector and supporting structures.

Frequency



Possibility of damaging the Product.

Frequency



Possibility of damaging the Product.

Frequency

**Hazard for the patient.**

Sterilization

6.4 Handpiece sterilization

The handpieces must be sterilized before use and can withstand up to 200 cycles.

The Operator must comply with the rules of the national commission for hygiene, disinfection and sterilization.

The handpieces are made of plastic material resistant to heat and knocks (PSU – Polysulfone).

Replace the handpieces as soon as these become cracked or deformed, as these could fall in the patient area.

Handpiece fitting / removal:

- Press the handpiece release lever and remove it.
 - Insert the handpiece up tight on the support and turn it until the steel lever engages in its original place and rotation is blocked.
- Finally make sure the handpiece is well secured.

Clean and disinfect the handpieces in the traditional way before sterilization. They can be cleaned with a mid-alkaline detergent free of active chlorine. To disinfect the handpieces, we suggest using alcohol or aldehyde-based products. The disinfectants must be approved by the manufacturer for use on polysulfone (PSU). After disinfecting, rinse off the detergent residues with plenty of water.

The handpieces fit into a suitable sterilization pack (disposable sterilization pack, e.g., plastic/paper bags; single or double pack), before being sterilized.

The handpieces can withstand about 200 steam sterilization cycles in accordance with the following parameters:

- steam sterilization at 121°C and 1.3 bar for 25 to 30 minutes
- steam sterilization at 134°C and 2.3 bar for 4 minutes

Do not exceed a sterilization temperature of 134°C.

Strictly keep to the ISO 17665-1 standard.

When placing in the autoclave, make sure the open side of the handpieces is turned downwards. The handpieces must be free and not burdened by other material being sterilized.

Damaged handpieces must no longer be used.

7 Adjustment and maintenance

7.1 Swinging arm adjustment

The Product is sold already balanced and does not require further adjustment. In the event of the spring swinging arm becoming stiff or loose over time, mechanical intervention is possible by regulating the compression of the internal spring.

Manually move the cover (1) forwards. Fit a pin (2) with diameter of 8 mm in the holes of the ring nut and turn in the direction indicated by the arrows to increase/decrease the load on the spring.

If the swinging arm drops, this means the elastic force of the spring is insufficient:

- turn the lever downwards and load the spring.

If the swinging arm lifts up, this means the elastic force of the spring is too high:

- turn the lever upwards and release the spring.


At the end of adjustment, manually reposition the cover (1) in its original position.

The swinging movement of the arm can also be adjusted upwards.

The Product is sold with maximum set swinging movement. If the swinging movement is to be reduced upwards, manually move the cover (1) forwards and insert a pin (3) with max diameter 5mm in the second ring nut. By turning the pin downwards, the swinging movement can be reduced until it is in horizontal position.

The swinging movement downwards cannot instead be changed.

At the end of adjustment, manually reposition the cover (1) in its original position.



7.2 Clutches adjustment


7.2.1 E/N SERIES clutches adjustment

The brakes are set during installation. Like all the other mechanical parts, the brakes are also subject to wear.

If the Product does not remain in a stable position, the braking force will have to be adjusted by means of the brake screws.

Use a flathead screwdriver to increase the braking force, turning the screws (1) and (2) of the arm brake clockwise.

To increase the braking force at the head, turn the two brake screws (3 and 4) clockwise using a flathead screwdriver.




7.2.2 TRIS-LED clutches adjustment

Brakes are regulated during the installation. All the mechanical parts, breaks included, are subjected to the normal wear and tear. In case the light head does not keep the position, is necessary to adjust the breaks acting on the screws.

Using a flathead screwdriver, tighten clockwise the screws (1) and (2), in order to strengthen the arm breaks.

Tighten clockwise the screws (3) and (4) using an allen key, to strengthen the light head break.





Perform the Product electrical check.



Making any changes to this device is forbidden.



Interrupt the power supply before doing any maintenance jobs.






Check Product integrity.

7.3 Periodical checks to be performed on the Product

At the time of start up and after each maintenance job, perform electrical tests and jobs indicated in the IEC 62353 standard.

7.4 Routine maintenance

N.	Period	Action
1	Before using	Make sure there are no pieces or fragments of paint that could become detached and fall within the operating field. If there are any, remove them manually.
2	Before using	Make sure the light source protection screens are not damaged. If they are, contact the Customer Service.
3	Once a year	Check all the Product joints and make sure there are no noises or squeaks. If there are, lubricate the clutches involved with suitable grease for industrial use at a service temperature between -30°C and + 120°C, type OKS 470 or with similar properties.
4	Once a year	If the Product fails to maintain a regular position, adjust the clutches as indicated at points 7.1 and 7.2 (arm and clutches adjustment) .
5	Once a year (CEILING VERSION) 	Make sure the bar retention screws (1) are tightened properly. Also check the bar horizontal arm retention screws (4). If these are not properly fastened, adequately tighten. To access the nuts (1), slide off the aluminium ring (2) downwards by loosening the screws and bar cover (3).

6	Once a year (WALL VERSION) (only PENTALED 30E/30N and TRIS-LES) 	Make sure the wall retention screws (1) are tightened properly. If these are not properly fastened, adequately tighten.
7	Once a year (MOBILE VERSION) (only PENTALED 30E/30N and TRIS-LED) 	Make sure the stem retention screws (1) under the base are tightened properly. If these are not properly fastened, adequately tighten.



The Product must only be opened and repaired by the Technical Service Personnel for the fuse change. All other repairs to be done by the manufacturer.





Interrupt the power supply before doing any maintenance jobs.



7.5 Repairs

The only repair job with which the technical assistance personnel are charged is the fuse change.

To access the fuses in the ceiling version, open the bar cover as indicated in point 5 of paragraph 7.4.



To access the fuses in the wall version, remove the 4 screws (1) and the closing box (2).



Making any changes to this device is forbidden.

To access the fuses in the mobile version, lift the ring (1), remove the screws (2) and lift the cover (3).

Remove the screw fuse carrier (1) from the terminal board and replace the fuse (2) making sure it is replaced with another of the same type.

If necessary, RIMSA will provide the wiring diagrams, the list of component parts, the descriptions, setting instructions or other information to assist the technical assistance personnel in repairing the parts of the Product indicated as repairable by the technical assistance personnel.

If the above indications are not enough to solve the problem, contact the after-sales service.

7.6 Disposal after use

Comply with applicable laws on waste disposal. This product must not be disposed of in standard waste disposal bins. To avoid risks for the environment and health deriving from the dispersion of polluting substances in the environment, separate the various internal component parts such as iron, aluminium, plastic and electrical material, and dispose of these through authorized channels so as to ensure correct recycling.

Disposal at end-of-life

7.7 Spare parts list



Only original spare parts must be used.

Description	Order code
Sterilisable handpiece	Z200518
Electronic board E and N SERIES	Z300632
Electronic board TRIS-LED	Z300677
Membrane keyboard E SERIES	Z300227
Membrane keyboard N SERIES	Z300234
Membrane keyboard TRIS-LED	Z143057
Switch O/I (for wall and mobile version)	Z300016
Switching power supply E and N SERIES	Z400629
Switching power supply TRIS-LED	Z170178
Fuse T2AH 250V '5x20'	Z400195
Fuse T10AH 250V '5x20'	Z400217
Fuse T4AH 250V '5x20'	Z400215

8 Technical properties**8.1 E SERIES technical properties**

Technical details of light	PENTALED 30E	PENTALED 81
Illumination E_c at 1 m distance $\pm 10\%$ [Lux]	160,000	
Colour temperature double selection [K]	4,500 / 5,000	
Colour rendering index R_a [-]	96	95
R_9 [-]	>90	
Light range diameter d_{50} [mm]	130	140
Light range diameter d_{10} [mm]	220	260
Lighting depth $L1+L2$ [mm] at 60%	600	720
Lighting depth $L1+L2$ [mm] at 20%	950	1030
Max irradiation [W/m^2]	580	
Irradiation / Illumination [mW/m^2lx]	3.68	
Max irradiation in UV [W/m^2]	0.004	
Power connection details		
Primary alternate voltage [Volt ac]	100 – 240	
Frequency [Hz]	50/60	
Power input [VA]	60	110
Light source	n°30 LEDs	n°81 LEDs
Duration of LED diode light source [hr] (this figure can vary according to power peaks and operating frequency)	60,000	
Light intensity control [%]	20 – 100	

General data			
Regulation		REGULATION (EU) 2017/745	
Classification of Medical Device		Class I	
Standards		IEC 60601-2-41	
Essential performance	Distribution of minimum and adequate lighting (luminous flux emitted by the ME equipment does not vary by more than 20% during use and the colour temperature and the colour rendering index are stable and are within the range 3000K-6700K and 85-100, respectively; E _c value shall be ≥ 40,000 lux and ≤ 160,000 lux).		
	Limitation of energy in the operating field (UV-irradiance for wavelengths below 400 nm does not exceed 10 W/m ² and the total irradiance E _e in the lighted area does not exceed 1000 W/m ² at a distance of 1000 mm; E _c value shall be ≥ 40,000 lux and ≤ 160,000 lux; E _e /E _c ≤ 6 mV/m ² lx).		
Colour		RAL 9003	
IP degree of protection		IP20	
Operating conditions		Continuous operation	
Handpiece steam sterilization		121°C at 1.3bar from 25 to 30 minutes. 134°C at 2.3bar for 4 minutes.	
Mains power voltage insulation means		Outside the product (main switch) for ceiling versions. Main switch for mobile and wall versions.	
Dimensions			
Diameter of lamp body [cm]		40	63
Light emission surface [cm ²] (4500K – 5000K)		483 – 762	1,493 – 1,920
Weight of ceiling, double ceiling, wall, mobile, mobile battery surgical light [kg]		39, 63, 27, 47, 57	47, 79, /, /, /
Markings			
		In conformity with REGULATION (EU) 2017/745	
All technical light measurements are to be deemed with a tolerance of ±6% for metrological and manufacturing reasons			

8.2 N SERIES technical properties

Technical details of light	PENTALED 30N	PENTALED 63N
Illumination E_c at 1 m distance $\pm 10\%$ [Lux]	160,000	
Colour temperature double selection [K]	4,500 / 5,000	
Colour rendering index R_a [-]	96	
R_9 [-]	> 90	
Light range diameter d_{50} [mm]	130	140
Light range diameter d_{10} [mm]	220	260
Lighting depth $L1+L2$ [mm] at 60%	700	560
Lighting depth $L1+L2$ [mm] at 20%	1150	1080
Max irradiation [W/m ²]	580	
Irradiation / Illumination [mW/m ² lx]	3.68	
Max irradiation in UV [W/m ²]	0.004	0.002
Focus on handle	Yes	
Power connection details		
Primary alternate voltage [Volt ac]	100 – 240	
Frequency [Hz]	50 / 60	
Power input [VA]	60	70
Light source	n°30 LEDs	n°72 LEDs
Duration of LED diode light source [hr] (this figure can vary according to power peaks and operating frequency)	60,000	
Light intensity control [%]	20 – 100	

General data			
Regulation		REGULATION (EU) 2017/745	
Classification of Medical Device		Class I	
Standards		IEC 60601-2-41	
Essential performance	Distribution of minimum and adequate lighting (luminous flux emitted by the ME equipment does not vary by more than 20% during use and the colour temperature and the colour rendering index are stable and are within the range 3000K-6700K and 85-100, respectively; E _c value shall be ≥ 40,000 lux and ≤ 160,000 lux).		
	Limitation of energy in the operating field (UV-irradiance for wavelengths below 400 nm does not exceed 10 W/m ² and the total irradiance E _e in the lighted area does not exceed 1000 W/m ² at a distance of 1000 mm; E _c value shall be ≥ 40,000 lux and ≤ 160,000 lux; E _e /E _c ≤ 6 mV/m ² lx).		
Colour		RAL 9003	
IP degree of protection		IP20	
Operating conditions		Continuous operation	
Handpiece steam sterilization		121°C at 1.3bar from 25 to 30 minutes. 134°C at 2.3bar for 4 minutes.	
Mains power voltage insulation means		Outside the product (main switch) for ceiling versions. Main switch for mobile and wall versions.	
Dimensions			
Diameter of lamp body [cm]		40	63
Light emission surface [cm ²] (4500K – 5000K)		483 – 762	1,423 – 1,828
Weight of ceiling, double ceiling, wall, mobile, mobile battery surgical light [kg]		39, 63, 27, 47, 57	47, 79, /, /, /
Markings			
		In conformity with REGULATION (EU) 2017/745	
All technical light measurements are to be deemed with a tolerance of ±6% for metrological and manufacturing reasons			

8.3 TRIS-LED technical properties

Technical details of light	TRIS-LED
Illumination E_c at 1 m distance $\pm 10\%$ [Lux]	130.000
Colour temperature double selection [K]	4.200
Colour rendering index R_a [-]	94
R_9 [-]	94
Light range diameter d_{50} [mm]	150
Light range diameter d_{10} [mm]	280
Lighting depth $L1+L2$ [mm] at 60%	900
Lighting depth $L1+L2$ [mm] at 20%	1500
Max irradiation [W/m^2]	470
Irradiation / Illumination [mW/m^2lx]	3,56
Max irradiation in UV [W/m^2]	0,003
Power connection details	
Primary alternate voltage [Volt ac]	100 – 240
Frequency [Hz]	50/60
Power input [VA]	90
Light source	n°28 LEDs
Duration of LED diode light source [hr] (this figure can vary according to power peaks and operating frequency)	60.000
Light intensity control [%]	25 – 100

General data		
Regulation		REGULATION (EU) 2017/745
Classification of Medical Device		Class I
Standards		IEC 60601-2-41
Essential performance	Distribution of minimum and adequate lighting (luminous flux emitted by the ME equipment does not vary by more than 20% during use and the colour temperature and the colour rendering index are stable and are within the range 3000K-6700K and 85-100, respectively; E_c value shall be $\geq 40,000$ lux and $\leq 160,000$ lux).	
	Limitation of energy in the operating field (UV-irradiance for wavelengths below 400 nm does not exceed 10 W/m ² and the total irradiance E_e in the lighted area does not exceed 1000 W/m ² at a distance of 1000 mm; E_c value shall be $\geq 40,000$ lux and $\leq 160,000$ lux; $E_e/E_c \leq 6$ mV/m ² lx).	
Colour		RAL 9003
IP degree of protection		IP20
Operating conditions		Continuous operation
Handpiece steam sterilization		121°C at 1.3bar from 25 to 30 minutes. 134°C at 2.3bar for 4 minutes.
Mains power voltage insulation means		Outside the product (main switch) for ceiling versions. Main switch for mobile and wall versions.
Dimensions		
Diameter of lamp body [cm]		40
Light emission surface [cm ²] (4500K – 5000K)		483 – 762
Weight of ceiling, double ceiling, wall, mobile, mobile battery surgical light [kg]		39, 63, 27, 47, 57
Markings		
CE		In conformity with REGULATION (EU) 2017/745
All technical light measurements are to be deemed with a tolerance of $\pm 6\%$ for metrological and manufacturing reasons		

9 EU Declaration of Conformity

In accordance with Article 19 and Annex IV of REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL, of 5 April 2017, on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC

Manufacturer: **RIMSA P. LONGONI S.r.l.**

Address of registered place of business: Via Monterosa, 18/20/22 – 20831 SEREGNO (MB) – ITALY

Single registration number (SRN): IT-MF-000009224

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Device identification: MINOR SURGICAL LUMINAIRE

Basic UDI-DI	Device code	Device name
++B880PENTA30EPAM5	PENTA30EPA	LAMP MODEL PENTALED 30E WALL
++B880PENTA30EPIMM	PENTA30E PI	LAMP MODEL PENTALED 30E MOBILE STAND
++B880PENTA30ESONA	PENTA30E SO	LAMP MODEL PENTALED 30E CEILING
++B880PENTA30E30ETP	PENTA30E+30E	LAMP MODEL PENTALED 30E DOUBLE CEILING
++B880PENTA81SOCW	PENTA81SO	LAMP MODEL PENTALED 81 CEILING
++B880PENTA8130EHR	PENTA81+30E	LAMP MODEL PENTALED 81+30E
++B880PENTA81818N	PENTA81+81	LAMP MODEL PENTALED 81+81
++B880PENTA30NPANJ	PENTA30NPA	LAMP MODEL PENTALED 30N WALL
++B880PENTA30NPIP2	PENTA30N PI	LAMP MODEL PENTALED 30N MOBILE STAND
++B880PENTA30NSOPP	PENTA30N SO	LAMP MODEL PENTALED 30N CEILING
++B880PENTA30N30NW8	PENTA30N+30N	LAMP MODEL PENTALED 30N DOUBLE CEILING
++B880PENTA63NSORD	PENTA63NSO	LAMP MODEL PENTALED 63N CEILING
++B880PENTA63N30NYG	PENTA63N+30N	LAMP MODEL PENTALED 63N+30N
++B880PENTA63N63NZ8	PENTA63N+63N	LAMP MODEL PENTALED 63N+63N
++B880TRISPALEDVE	TRISPA-LED	LAMP MODEL TRIS-LED WALL
++B880TRISPILEDX6	TRISPI-LED	LAMP MODEL TRIS-LED MOBILE STAND
++B880TRISSOLEDZH	TRISSO-LED	LAMP MODEL TRIS-LED CEILING
++B880TRISSOX2LEDAW	TRISSOX2-LED	LAMP MODEL TRIS-LED DOUBLE CEILING

Risk class of the device in accordance with the rules set out in Annex VIII of REGULATION (EU) 2017/745: **CLASS I**

Explanation: Duration: Short term (Annex VIII, CHAPTER I, point 1. DURATION OF USE)

Description: Non-invasive medical device (Annex VIII, CHAPTER III, point 4. NON-INVASIVE DEVICES, par. 4.1 Rule 1)

Active medical device (Annex VIII, CHAPTER III, point 6. ACTIVE DEVICES, par. 6.2 Rule 10)

The manufacturer declares that the medical device is in conformity with REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL, of 5 April 2017, on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC and with the following standards:

- IEC 60601-1 (Part 1: General requirements for basic safety and essential performance)
- IEC 60601-1-2 (Part 2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests)
- IEC 60601-2-41 (Part 1: Particular requirements for the safety of surgical luminaires and luminaires for diagnosis)

The conformity assessment procedure is developed with reference to premise (60) and Article 52 of REGULATION (EU) 2017/745.

RIMSA Quality System complies with UNI EN ISO 9001 and UNI CEI EN ISO 13485 standards and is certified by CSQ (CSQ certificate no. 9120.RMS1 and 9124.RMS2).

Name: Paolo Longoni
Position: Managing Director





Possibility of interferences with nearby appliances.

10 EMC Declaration

The Product has been tested according to IEC 60601-1-2 standard to ensure correct electromagnetic compatibility.

Portable and mobile communication appliances can affect the Product. The product should not be used close to another device and if this is inevitable, the product must be checked to make sure it is working properly.

The use of accessories other than those supplied/recommended by the manufacturer could increase the level of emissions and lower the level of immunity of the appliance.

The Product has been designed to be used in the electromagnetic environments described below.

The Responsible Organization or Operator is responsible for making sure the Product is used in a compatible environment.

It could occur that if the Product is affected by radiations in the range of 80 MHz – 1 GHz or bursts, it will no longer respond to the commands both as regards the lamp and the camera.

If this does occur, essential performance will in any case be ensured, but to restore normal operation it will be necessary to de-energize the master switch.

Immunity test	Compliance	Electromagnetic environment – directives
RF Emissions CISPR 11	Group 1	The Product only uses RF energy for internal operation. Consequently its RF emissions are very low and should not cause any interference to nearby electronic appliances.
RF Emissions CISPR 11	Class A	The Product is suitable for use in all environments except in domestic environments and those directly connected to a low-voltage public mains supply which supplies buildings used for domestic purposes, as long as the following precaution is followed. Warning: This Product is intended for use by professional health personnel only. This Product can cause radio-interference or disturb the operation of nearby appliances. Measures may have to be taken to reduce such disturbance, such as Product re-positioning or shielding of premises.
Harmonic emissions IEC 61000–3–2	Class A	
Voltage fluctuations /flicker emissions IEC 61000–3–3	Conforming	
NOTE: The EMISSIONS characteristics of this equipment make it suitable for use in industrial areas and hospitals (CISPR 11 class A). If it is used in a residential environment (for which CISPR 11 class B is normally required) this equipment might not offer adequate protection to radio-frequency communication services. The user might need to take mitigation measures, such as relocating or re-orienting the equipment.		

Immunity test	Test level to IEC 60601-1-2	Conformity level	Electromagnetic environment - directives
Electrostatic discharge (ESD) IEC 61000-4-2	+/- 8 kV at contact +/- 15 kV in air	+/- 8 kV at contact +/- 15 kV in air	Floors must be made of wood, concrete or ceramic tiles. If the floors are covered with synthetic material, relative humidity must at least be equal to 30%.
Rapid impulse electric transistors IEC 61000-4-4	+/- 2 kV For electric power lines +/- 1 kV For input/output lines	+/- 2 kV For electric power lines +/- 1 kV For input/output lines	Mains voltage quality should be that of a typical commercial or hospital environment.
Overvoltage IEC 61000-4-5	+/- 1 kV Between phases +/- 2 kV Between phases and earth	+/- 1 kV Between phases +/- 2 kV Between phases and earth	Mains voltage quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and variations on the power supply input lines IEC 61000-4-11	<5% U_T (drop >95% of U_T) For 0.5 cycles 40% U_T (drop = 60% of U_T) For 5 cycles 70% U_T (drop = 30% of U_T) For 25 cycles <5% U_T (drop >95% of U_T) For 5 s	<5% U_T (drop >95% of U_T) For 0.5 cycles 40% U_T (drop = 60% of U_T) For 5 cycles 70% U_T (drop = 30% of U_T) For 25 cycles <5% U_T (drop >95% of U_T) For 5 s	Mains voltage quality should be that of a typical commercial or hospital environment. If the Product user requires continued function during mains power supply interruptions, the Product should be supplied by a UPS unit or batteries.
Magnetic field at electrical mains frequency (50/60Hz) IEC 61000-4-8	30 A/m	30 A/m	The magnetic fields at mains frequency should have the characteristic levels of a typical locality in a commercial or hospital environment.

NOTE: U_T mains voltage in AC before application of test level.

Immunity test	Test level to IEC 60601-1-2	Conformity level	Electromagnetic environment - directives
Conducted RF IEC 61000-4-6	3 Veff 150 kHz to 80 MHz	3 Veff	Portable and mobile RF communications equipment should be used no closer to any part of the Products, included cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance: $d = 1.2\sqrt{P}$ 150 KHz to 80 MHz $d = 1.2\sqrt{P}$ 80 MHz to 800 MHz $d = 2.3\sqrt{P}$ 800 MHz to 2.7 GHz where P is the maximum output power rating of the transmitter in watts (W), according to the transmitter manufacture and d is the recommended separation distance in meters (m). Field strengths from fixed transmitters, as determined by an electromagnetic site survey, should be less than the compliance leave in each frequency range. Interference may occur in the vicinity of equipment marked with the following symbol: 
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.7 GHz	3 V/m	

NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Test frequency (MHz)	Band ^{a)} (MHz)	Service ^{a)}	Modulation ^{b)}	Maximum power (W)	Distance (m)	IMMUNITY TEST LEVEL (V/m)
385	380-390	TETRA 400	Pulse modulation ^{b)} 18 Hz	1.8	0.3	27
450	430-470	GMRS 460, FRS 460	FM ^{c)} ± 5kHz deviation 1 kHz sine	2	0.3	28
710	704-787	LTE Band 13, 17	Pulse modulation ^{b)}	0.2	0.3	9
745						
780			217 Hz			
810	800-960	GSM800/900, TETRA 800, iDEN 820, CDMA 850, LTE Band 5	Pulse modulation ^{b)}	2	0.3	28
870						
930			18 Hz			
1720	1700-1990	GSM 1800; CDMA 1900; GSM 1900; DECT; LTE Band 1, 3, 4, 25; UMTS	Pulse modulation ^{b)}	2	0.3	28
1845						
1970			217 Hz			
2450	2400-2570	Bluetooth, WLAN, 802.11 b/g/n, RFID 2450, LTE Band 7	Pulse modulation ^{b)} 217 Hz	2	0.3	28
5240	5100-5800	WLAN 802-11 a/n	Pulse modulation ^{b)}	0.2	0.3	9
5500						
5785			217 Hz			

NOTE: If necessary to achieve the IMMUNITY TEST LEVEL, the distance between the transmitting antenna and the ME EQUIPMENT or ME SYSTEM may be reduced to 1 m. the 1m test distance is permitted by IEC 61000-4-3.

a) For some services, only the uplink frequencies are included.

b) The carrier shall be modulated using a 50% duty cycle square wave signal.

c) As an alternative to FM modulation, 50% pulse modulation at 18 Hz may be used because while it does not represent actual modulation, it would be worst case.

Recommended separation distance between portable and mobile RF communications equipment and the Product

The Product is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the Product can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Product as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter m		
	150 kHz to 80 MHz $d = 1.2\sqrt{P}$	80 MHz to 800 MHz $d = 1.2\sqrt{P}$	800 MHz to 2.7 GHz $d = 2.3\sqrt{P}$
0.01	0.12	0.12	0.24
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

11 Warranty Certificate

1. The Product is covered by an 18-month warranty, including electrical parts
2. The warranty begins on the date of Product shipment from the RIMSA warehouse to the buyer.
3. In case of disputes, the date indicated on the "transport document" attached to the goods shall be deemed valid.
4. The warranty only covers the sending of Product spare parts to the buyer or, in the event of RIMSA considering the replacement of spare parts not feasible, the replacement of the entire product, after fabrication faults have been properly ascertained at the undisputable judgement of RIMSA. The warranty does not therefore cover any other costs or expenses (including, by way of example but without limitation, labour costs, packaging costs and transport costs, etc.).
5. The guarantee does not include the components subject to normal wear, such as halogen bulbs, LEDs, fuses, relays, ball bearings, etc.)
6. The warranty does not cover:
 - malfunctions due to failure to comply with all instruction manuals;
 - malfunctions due to installation and/or maintenance errors;
 - malfunctions or faults caused by carelessness, negligence, incorrect use or other causes not attributable to RIMSA;
 - malfunctions or faults due to the fact that the electrical system of the premises where the device is installed is not in compliance with IEC 60364-7-710 standard (standard for electrical systems in premises used for medical purposes) and similar standards.
7. RIMSA shall repay direct damages suffered by the buyer and which are documented as attributable to its product, caused within the warranty period, for an amount not above 40% of the net value of the product as indicated on the buyer's invoice. RIMSA's liability is expressly ruled out for indirect damages or consequential damages (including cases of the Product not being used) deriving from the supply.
8. This warranty certificate replaces legal warranties for faults and non-conformities and rules out any other possible liability of RIMSA originating from the supplied products.
9. The payment of any damages to persons or things due to product malfunction or faults shall be limited to the maximum amount of RIMSA's insurance coverage for civil liability.
10. The warranty shall be automatically invalidated in the event of:
 - the Product having been tampered with or modified by the buyer or third parties;
 - the Product having been repaired by the buyer or third parties, without following the instructions in the instruction manuals;
 - the Product serial number having been cancelled, defaced or removed;
 - the buyer not being up to date with payments.
11. For jobs to be done under warranty, the buyer shall contact RIMSA only.
12. The component parts replaced under warranty must only be returned to RIMSA, if so requested by RIMSA, carriage free and suitably packed.
13. In case of failure to return a part requested by RIMSA, the cost of the component part will be charged.
14. RIMSA cannot accept returns from end users or in any case from parties other than the buyer.
15. Products returned to RIMSA must be complete with documentation authorising such return and another document describing the malfunction.
16. For everything not indicated on this warranty certificate, reference shall be made to the laws of Italy
17. For all disputes deriving from or related to the orders to which this warranty certificate applies and which cannot be amicably settled between the parties, the only competent law court shall be that of Milan.

Notes