

DEAMBULATORE A 3 RUOTE ROLLATOR WITH 3 WHEELS DÉAMBULATEUR 3 ROUES ROLLATOR MED 3 HJUL



All serious accidents concerning the medical device supplied by us must be reported to the manufacturer and competent authority of the member state where your registered office is located.

INTRODUCTION

This manual provides the instructions for a correct use of our patient aid or an assistant / carer.

INTENDED USE

The rollator is designed to be used inside buildings by patients with an impairment

The devices are intended to provide a support to the body. This aid allow patient mobility making it easy to walk. They must be used with two arms in order to guarantee safe operation.

WARNING! The rollator is built for indoor use. If it is necessary to employ the devices outside we advise to follow the road rules like any pedestrian.

Before each use make sure that:

- The walker has been adjusted to the correct height.
 The tips are in good condition. Should they be worn or missing, replace them immediately.
- · Make sure that the legs of the walker have been adjusted to the same height.

MAINTENANCE

The rollator do not require special care for their ordinary maintenance. We advise to regularly perform the following checks:

- verify the reliability of different components
- check the correct fi xing of screws and nuts
- test moving parts, apply lubricant if needed
 clean parts with mild detergents (do not use corrosive or harmful chemicals), put it in dry and ventilated environment.


WARNING!

Do not use the equipment in case it is damaged or any component is missing. Do not use the patient aids for purposes not mentioned in this user manual. Do not cross obstacles or use the patient aids on uneven surfaces, or wet and slippery. The use of the device is not recommended for people with poor functionality in the hands and arms.


Only use original accessories and spare parts. The use of non-original accessories or spare parts could cause the risk of falls and injuries. Do not use on stairs. The use of walkers in water (showers, bathtubs, swimming pools, etc.) can favour the oxidation of the operating mechanisms, causing damage to the walker and/or injury to the user.

Max load 100 kg

ASSEMBLY INSTRUCTION



Open the folded "frame assembly" in the direction shown in the figure



Step 2: Hang the basket on the hooks



Step 3: Put the handle pipe assembly around the two sides of the

frame socket, and adjust the height to the appropriate height to lock the nut. Try to push the rollator, check the reliability of the brake, adjust it properly according to the characteristics of each person.

	Manufacturer	*	Keep in a cool, dry place
[]i	Consult instructions for use	淤	Keep away from sunlight
REF	Product code	LOT	Lot number
سا	Date of manufacture	MD	Medical Device
EC REP	Authorized representative in the European community	\triangle	Caution: read instructions (warnings) carefully
CE	Medical Device compliant with Regulation (EU) 2017/745		

GIMA WARRANTY TERMS

The Gima 12-month standard B2B warranty applies.



FS969H (43140)



Foshan Dongfang Medical Equipment Manufactory (LTD.) 5 Xingfu Road, Zone A, Shishan Industrial Park, Nanhai District, Foshan City, China, 528000 Made in China

Importato da / Imported by / Importé par / Importerad av:

Gima S.p.A. Via Marconi, 1 - 20060 Gessate (MI) Italy gima@gimaitaly.com - export@gimaitaly.com www.gimaitaly.com



MedNet EC-Rep GmbH Borkstrasse 10. 48163 Muenster. Germany











