



Product Code	29904
Unit of sale	1 pc
Minimum order	1
Type	Medical device
Class	II B

UK-REP	Yes
--------	-----

RDM (NSIS)	1295791
CND	U089005
EAN/UPC	8055060761677

Description

Blue silicone pessary, for uterine prolapse.

- Sterile
- Latex free
- Autoclavable at 121°C/15 min or 134°C/10 min for up to 3 times

Individually packed in cardboard box. Multilingual box and instructions: GB, FR, IT, ES, PT, DE, GR.