

文件名称: C-0852-20-77435 机器标贴 (英文CE)

作者: 李琼 04915

文件编号: C-0852-20-77435


文件类型: 技术文件

生效日期: 06/18/2015 10:10:25

版本: 4.0

审核	赵燕	50215353	赵燕	05/21/2015 18:39:06
审核	贾智娱	50219718	贾智娱	05/21/2015 17:46:05
审核	杨义鹏	50003488	杨义鹏	05/21/2015 09:31:14
审核	邹勇	50007153	邹勇	05/20/2015 17:39:19
审核	陈明华	50007823	陈明华	05/20/2015 15:23:41
审核	石浪涛	50211069	石浪涛	05/21/2015 08:47:54
文控审核	刘榕枝	50213456	刘榕枝	05/22/2015 13:40:17
批准	刘继广	50003438	刘继广	05/22/2015 14:05:20

REV	ECN/PCN/CR	DESCRIPTION	DESIGN
3.0	EF490	删除电击防护标识，换用新图框等	丁晓亮
4.0	ECG009	将 ⚠️ 更改为 🏥	李琼



技术要求

- 1、在光滑面丝印，光滑面为反面，文字与符号的大小、位置等依1：1的电子文档；
- 2、颜色要求：文字、符号与线框颜色为Panton432C，底色为MR-C01-S（白色）；
- 3、图上未注尺寸公差±0.3mm；
- 4、丝印面背3M 468MP胶，背胶要求粘贴力强，耐湿；
- 4、来料时在外包装上贴上物料标贴。

APPROVALS		DATE	mindray		保密密级：秘密
DESIGN	杨义鹏		TITLE		SECRET
CHECK			机器标贴 (英文CE)		
CHECK			DWG NO.	C-0852-20-77435	
CHECK			MATERIAL	8B35 PC T=0.125	P.CODE 0852
APPROVAL			SCALE	1 / 1	REV. 4.0
Software & Rev: CoreIDRAW X3			PROJECTION	SIZE A4	SHEET 1 OF 1

保密：此图及其全部知识产权（含著作权）归深圳迈瑞生物医疗电子股份有限公司所有。未经深圳迈瑞生物医疗电子股份有限公司预先书面许可，严禁出于任何目的，对此图的全部或部分内容（包括但不限于图中信息、数据、运算结果等）泄露、使用、拷贝或复制。

Classified documents. This set of drawing(s) and all its intellectual property rights (including copyright) subsisting herein are property of Shenzhen Mindray Bio-medical Electronics Co., Ltd. No disclosure, use, copies or reproductions should be made of this drawing or any part(s) thereof for whatever purpose nor shall any information, data, calculations, or other contents contained in this drawing be disseminated without prior written permission of Shenzhen Mindray Bio-medical Electronics Co., Ltd.