



山东连发医用塑胶制品有限公司
Shandong Lianfa Medical Plastic Products Co.,Ltd

EU Declaration of Conformity

Manufacturer:

Shandong Lianfa Medical Plastic Products Co., Ltd.
No.1 Shuangshan Sanjian Road, Zhangqiu , Jinan
City,250200 Shandong P. R. China
SRN: CN-MF-000028790

Whose single Authorized Representative:

Linkfar Healthcare GmbH
Niederrheinstraße 71, 40474 Düsseldorf, Germany
SRN: DE-AR-000005107

We, the manufacturer, herewith declare that the products

Product Name:Lancing Devices

Model: LDA; LDC; LDE; LDF; LDC-II; LDE-II; LDF-II

EMDN Code:V9099

Basic UDI-DI: 694951701VA

Intended use:

Assisting the lancet for peripheral blood sampling from the fingertip or the paw for blood glucose testing or other testing utilising a small amount of blood, it is the non-invasive device and does not come into contact with blood, it only makes a brief contact with the skin surface for a few seconds. The lancing device is non-sterile.

Standards applied:

EN ISO 20417:2021, EN ISO 15223-1:2021,EN ISO 10993-1:2020, EN ISO 10993-5:2009, EN ISO 10993-10:2023, EN ISO 10993-23:2021;EN ISO 14971:2019, EN 62366-1:2015, ASTM F 1980:2016, ASTM D 4169:2016, EN ISO 13485:2016

Meet the provisions of Regulation (EU) 2017/745 which apply to them.

The medical device has been assigned to class I according to Rule 1, Annex VIII of the Regulation (EU) 2017/745. It bears the mark



The above mentioned declaration of conformity is exclusively under the responsibility of

Company: Shandong Lianfa Medical Plastic Products Co., Ltd.

Address: No.1Shuangshan SanjianRoad, Zhangqiu,Jinan City,250200 Shandong P. R. China

Jinan 2025-1-11

Place, date

Lianying Yang|CEO

Legally binding signature, Function

