


KANEHARA TRADING INC.
31-14,2-chome Yushima Bunkyo,
Tokyo 113-0034 JAPAN
TEL:81-3-3813-0287
FAX:81-3-3813-0285

EU-Declaration of Conformity

We, KANEHARA TRADING INC, declare on our own responsibility that the medical devices described below are in compliance with requirements of the following regulation and their related harmonized standards: (EU) 2017/745

Manufacturer's name	KANEHARA TRADING INC. 31-14,2-chome Yushima Bunkyo, Tokyo 113-0034, JAPAN SRN: TBD
European Representative	OCULUS Optikgeräte GmbH Münchholzhäuser Str. 29 D-35582 Wetzlar, Germany SRN: TBD
Product name	Ishihara's Tests for Color Deficiency
Article number and type	47600 Ishihara's Tests for Color Deficiency (24 plates) 47610 Ishihara's Tests for Color Deficiency (14 plates) 47620 Ishihara's Tests for Color Deficiency (10 plates) 47640 Ishihara's Tests for Color Deficiency (38 plates)
Basic UDI-DI	4580759476AS
Conformity Assessment Procedures	Gemäß: EU/2017/745: Artikel 19 / Anhang IV according: EU/2017/745, article 19 / Annex IV
Applied Common Specifications	N/A
Risk class	I
Classification rule	1
This declaration of conformity is valid until	10.10.2030.

Place, date	Name and function
Tokyo, 21.10.2025	 Name TOSHIE KANEHARA Position President