DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

CONTEC MEDICAL SYSTEMS CO., LTD

No.112 Qinhuang West Street, Economic & Technical MANUFACTURER:

Development Zone, Qinhuangdao, Hebei Province,

PEOPLE'S REPUBLIC OF CHINA

MEDICAL DEVICE: Ambulatory Blood Pressure Monitor, ABPM50

CLASSIFICATION - ANNEX IX: Class II a, Rule 10

CONFORMITY ASSESSMENT ROUTE: Annex II excluding chapter 4

WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES MEET THE TRANSPOSITION INTO NATIONAL LAW. THE PROVISIONS OF COUNCIL

DIRECTIVE 93/42/EEC of 14 JUNE 1993 CONCERNING MEDICAL DEVICES;

INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.

THIS EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE RESPONSIBILITY OF THE MANUFACTURER.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

TÜV SÜD PRODUCT SERVICE GMBH NOTIFIED BODY:

RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

((0123 **IDENTIFICATION NUMBER:**

G1 050972 0050 Rev.04 (EC) CERTIFICATE(S):

Prolinx GmbH

EUROPEAN REPRESENTATIVE: Brehmstr. 56, 40239, Duesseldorf, Germany

PLACE, DATE OF DECLARATION: QINHUANGDAO, 2024/01/26

SIGNATURE:

President

TF-CE100302-09 Ver:L Page 1 of 2

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

NO.	Standards	Title and Description
1	ISO 13485:2016	Medical devices – Quality management systems –
		Requirements for regulatory purposes
2	ISO 14971:2019	Medical devices – Application of risk management to medical
		devices
3	IEC 60601-1:2005+AMD1:2012	Medical electrical equipment- Part1: General requirements for
		basic safety and essential performance
4	IEC 60601-1-2:2014	Medical electrical equipment- Part 1-2: General requirements
		for basic safety and essential performance - Collateral
		standard: Electromagnetic disturbances - Requirements and
		tests
5	IEC60601-1-6:2010+AMD1:2013+AMD2: 2020	Medical electrical equipment - Part 1-6: General requirements
		for basic safety and essential performance - Collateral
		standard: Usability
6	IEC 62366-1:2015+AMD1:2020	Medical devices - Part 1: Application of usability engineering
		to medical devices
7	IEC 62304:2006+AMD1:2015	Medical device software-Software life-cycle processes
8	EN ISO 15223-1:2021	Medical devices - Symbols to be used with information to be
		supplied by the manufacturer - Part 1: General requirements
9	ISO 20417:2021	Medical devices-Information to be supplied by the
		manufacturer
10	IEC 80601-2-30:2018	Medical electrical equipment Part 2: Particular requirements
		for the basic safety and essential performance of automated
		non-invasive sphygmomanometers
11	ISO 10993-1:2018	Biological evaluation of medical devices - Part 1: Evaluation
		and testing within a risk management process
12	IEC 60601-1-11:2015	Medical electrical equipment-Part 1-11: General requirements
		for basic safety and essential performance- Collateral
		Standard: Requirements for medical electrical equipment and
		medical electrical systems used in the home healthcare
		environment

TF-CE100302-09	Ver:L	
Page 2 of 2		