

METRAX

GmbH Rheinwaldstr. 22
D-78628 Rottweil

Made in Germany

Type
M290

HIBC



+EPRI735030/\$\$+123473503000000T



F.-Code
1234 SN 735030000000

PRIMEDIC
Defi Monitor XD1

100-240VAC 50/60Hz 129W
(short term) / 12VDC / 7,5A

Operation only with (rechargeable) battery inserted

C **E** 0123



ECG

IP33

