

DICHIARAZIONE NON DISPOSITIVO MEDICO**DECLARATION NOT MEDICAL DEVICE**

I seguenti prodotti non sono classificati come dispositivi medici, pertanto non è richiesta la Dichiarazione di Conformità ai sensi del Regolamento (UE) 2017/745.

Following products are not classified as medical devices, therefore Declaration of Conformity as per Regulation (EU) 2017/745 is not required.

Descrizione prodotto <i>Product description</i>	Riferimento # \ CODICE PRODOTTO <i>Reference #</i> \PRODUCT CODE	Destinazione d'uso <i>Intended use</i>
DROP SENSOR for IP-7700	IPDS013	Checking drip status
DOSAGE AND PROFILE PROGRAM	DPV01	Operation program of infusion condition

Infatti, la destinazione d'uso del prodotto non è compatibile con la definizione di dispositivo medico.

In fact, the intended use of the product doesn't match with the definition of medical device.

Nome del Fabbricante/ <i>Manufacturer name</i>	Indirizzo postale ed elettronico a cui il fabbricante può essere contattato / <i>postal and electronic address at which manufacturer can be contacted</i>
AMPALL CO., LTD.	3F Annex Hankook Junja Hyeopdong B/D 114 Gasan Digital 2-ro, Geumcheon-gu, Seoul, Korea www.ampall.com / ampall@ampall.com

Luogo, data/*Place, date* : SEOUL, REPUBLIC OF KOREA

April 11, 2025

Deukyoung Song / President

Nome, Firma, funzione, timbro

Name, Signature, job position, stamp

AMPALL CO., LTD.


Deukyoung Song / President