



DECLARATION OF CONFORMITY

PVS SpA located in Leonardo da Vinci, 18 street , 20051 Cassina De' Pecchi (Mi), Italy, as assembler of the following kit :

Description	MIZARD KIT
REF	CPS064
Basic UDI-DI	80340280114D SRN - IT-PR-000050312

Composition :


- 2 Pairs of sterile gloves; MD class Is
- 1 Disinfectant bottle 125 ml IODOPOVIDONE
- 1 sterile solution 250 ml MD class IIa
- 1 Sterile gauze compressed 18×40 cm; MD class Is
- 3 Sterile gauze compressed 10×10 cm; MD class Is
- 1 Sterile tweezers ; MD class Is
- 1 Pack of cotton wool; MD class I
- 1 PLASTOSAN 10 assorted plasters; MD class I
- 1 Adhesive plaster spool m 5×2.5 cm; MD class I
- 1 Gauze bandage 4x10 cm; MD class I
- 1 Pair cutting scissors; MD class I
- 1 Tourniquet; MD class I
- 1 ICE PACK ; MD class IIa
- 1 Bag for medical waste 250×350 mm.

In compliance with Article 22 of REG. UE 2017/745, declares that :

- a) the mutual compatibility of the devices has been verified;
- b) is responsible for assembly, packaging and has provided users with the relevant information containing the relevant manufacturer's instructions
- (c) the activity of combining the devices into procedural kits has been subjected to adequate methods of internal control, verification and validation.
- d) the system incorporates both CE medical devices and non-medical devices.

Cassina De' Pecchi, 30.09.2025

LEGAL REPRESENTATIVE


IRENE PEREGO