



Equipment Under Test (EUT)	: Digital Video Colposcope
Model Number of EUT	: COLpro222DX-OZview
Serial Number of EUT	: DVC010-0510
Submitted by	: M/s. Frontline Electro Medical Ltd., New Delhi



**EMI/EMC TEST REPORT FOR DIGITAL VIDEO COLPOSCOPE
SUBMITTED BY M/S. FRONTLINE ELECTRO MEDICAL LTD., NEW DELHI**


Test Request Particulars


1. Test Request From : M/s. Frontline Electro Medical Ltd., New Delhi
2. Equipment Under Test (EUT) : Digital Video Colposcope
3. Number of Test Sample(s) : One
4. Types of Tests Requested
(Applicable Standard)
 1. Conducted Emission Test
 2. Radiated Emission Test
 3. Harmonic Current Emission Test
 4. Flicker Emission Test
 5. Radiated Susceptibility Test
 6. Electrostatic Discharge Immunity Test
 7. Electrical Fast Transient Immunity Test
 8. High Energy Surge Immunity Test
 9. Conducted RF Immunity Test
 10. Power Frequency Magnetic Field Test
 11. Voltage Dips and Interruption Test- All tests as per IEC 60601-1-2: 2007
5. Manufacturer by : M/s. Promis Medical (Australia) Pty Ltd., Australia
6. Model Number of EUT : COLpro222DX-OZview
7. Serial Number of EUT : DVC010-0510
8. Test plan concurred by
(Customer Representative) : Mr. B. Narasimha Murthy, Sales and Service Engineer
Frontline Electro Medical Ltd., New Delhi
9. EUT Arrived On : January 19, 2011
10. Tested On : January 19-21, 2011
11. Test Venue : SAMEER-CEM, Chennai
12. Status of the EUT on Receipt : Functional

Certified that the data reported in this report are valid only for the test sample mentioned above at the time of and under the stated conditions of measurement. Particulars on Manufacturer / Supplier, given in this report, are based on the information given by the customer, along with test request and SAMEER-CEM does not assume any responsibility for the correctness of that information for the above mentioned equipment under test.

Test Plan & Reviewed by:

Approved by:


(P. Salil)
Scientist-D


(Dr. B. Subbarao)
Head, EMC Division



S-CEM/EMCD/TR/2010-2011/367
Page 2 of 35